

Integrating the Healthcare Enterprise



5

**IHE Cardiology
Technical Framework Supplement**

10

**Cardiac Imaging Report Content
(CIRC)**

15

Trial Implementation

Date: July 1, 2011
20 Editor: IHE Cardiology Technical Committee 2010-2011
Email: cardio@ihe.net

Foreword

25 This is a supplement to the IHE Cardiology Technical Framework. Each supplement undergoes a process of public comment and trial implementation before being incorporated into the volumes of the Technical Frameworks.

This supplement is submitted for Trial Implementation as of July 1, 2011 and will be available for testing at subsequent IHE Connectathons. The supplement may be amended based on the results of testing. Following successful testing it will be incorporated into the Cardiology Technical Framework. Comments are invited and may be submitted on the IHE forums at 30 <http://forums.rsna.org/forumdisplay.php?f=249> or by email to cardio@ihe.net.

This supplement describes changes to the existing technical framework documents and where indicated amends text by addition (**bold underline**) or removal (~~**bold strikethrough**~~), as well as addition of large new sections introduced by editor’s instructions to “add new text” or similar, which for readability are not bolded or underlined.

35 “Boxed” instructions like the sample below indicate to the Volume Editor how to integrate the relevant section(s) into the relevant Technical Framework volume:

<i>Replace Section X.X by the following:</i>
--

40 General information about IHE can be found at: www.ihe.net

Information about the IHE Cardiology can be found at: <http://www.ihe.net/Domains/index.cfm>

Information about the structure of IHE Technical Frameworks and Supplements can be found at: <http://www.ihe.net/About/process.cfm> and <http://www.ihe.net/profiles/index.cfm>

45 The current version of the IHE Technical Framework can be found at: http://www.ihe.net/Technical_Framework/index.cfm

CONTENTS

	INTRODUCTION	7
50	RELATIONSHIP TO WORKFLOW PROFILES	8
	OPEN ISSUES AND QUESTIONS	8
	CLOSED ISSUES	9
	VOLUME 1 – INTEGRATION PROFILES	12
	1.7 CURRENT YEAR SCOPE.....	12
55	2.1 DEPENDENCIES AMONG INTEGRATION PROFILES	12
	2.2.11 <i>Cardiac Imaging Report Content (CIRC)</i>	12
	2.3 ACTOR DESCRIPTIONS	13
	11 CARDIAC IMAGING REPORT CONTENT (CIRC) INTEGRATION PROFILE	13
	11.1 ACTORS/ TRANSACTIONS.....	14
60	11.1.1 <i>Actor Groupings</i>	15
	11.2 CARDIAC IMAGING REPORT CONTENT PROFILE OPTIONS	15
	11.2.1 <i>Content Consumer Options</i>	15
	11.3 USE CASES.....	15
	11.3.1 <i>Compile and transfer report of cardiac imaging procedure with use of key imaging data measures</i>	15
65	11.3.2 <i>Perform discrete data-analysis on procedure report content</i>	16
	11.3.3 <i>Review Procedure Report</i>	17
	11.4 SECURITY CONSIDERATIONS.....	17
	11.5 GROUPING	17
	11.5.1 <i>Content Bindings for Displayable Reports (DRPT) and Image Enabled Office (IEO) Profiles</i>	18
70	11.5.2 <i>Content Bindings for XDS, XDM, XDR, and XDS-I</i>	18
	11.5.3 <i>Content Binding for Portable Data for Imaging (PDI)</i>	18
	11.5.4 <i>Content Binding for Retrieve Form for Data Capture (RFD)</i>	18
	11.5.5 <i>Relationship to Document Digital Signature (DSG)</i>	19
	11.6 REQUIREMENTS OF CIRC ACTORS	19
75	11.6.1 <i>Content Creator</i>	19
	11.6.2 <i>Content Consumer</i>	19
	VOLUME 2 – TRANSACTIONS AND CONTENT MODULES	21
	3 FRAMEWORK OVERVIEW.....	21
	3.1 ACTORS AND TRANSACTIONS	21
80	3.2 CONTENT MODULES.....	21
	5 NAMESPACES AND VOCABULARIES.....	22
	5.1 IHE FORMAT CODES	22
	6 CONTENT MODULES	24
	6.1 CONVENTIONS	24
85	6.1.1 <i>Content Module Conventions</i>	24
	6.1.1.1 <i>Cardinality Constraints</i>	24
	6.1.1.2 <i>Data Element Optionality Constraints</i>	24
	6.1.1.3 <i>Coded Terminology Values</i>	25
	6.1.2 <i>Structure of Content Modules</i>	26
90	6.1.2.1 <i>Document Content Modules</i>	27
	6.1.2.1.1 <i>Document Content Module Table</i>	27
	6.1.2.2 <i>Section Content Modules</i>	28
	6.1.2.2.1 <i>Section Content Module Table</i>	29

	6.1.2.2.2 Observation Entry Constraint Table.....	30
95	6.1.2.3 Entry and Header Content Modules	31
	6.1.2.3.1 Header Content Module Table	31
	6.1.2.3.2 Entry Content Module Table.....	32
	6.1.2.4 Value Sets	33
	6.2 CDA RELEASE 2 CONTENT MODULES.....	34
100	6.2.1 Document Modules	34
	6.2.1.1 Cardiac Imaging Report Specification 1.3.6.1.4.1.19376.1.4.1.1.1	34
	6.2.1.1.1 Format Code	34
	6.2.1.1.2 Parent Template	34
	6.2.1.1.3 Clinical Document Code.....	34
105	6.2.1.1.4 Standards.....	34
	6.2.1.1.5 Data Element Requirements.....	35
	6.2.1.1.6 Specification	37
	6.2.1.1.6.1 Service Event Constraints	39
	6.2.1.1.6.2 Medications Section Constraints.....	39
110	6.2.1.1.6.3 Allergies and Other Adverse Reactions Section Constraints	39
	6.2.1.1.6.4 Coded Social History Section Constraints	39
	6.2.1.1.6.5 Indications and Planned Procedure Section Constraints	39
	6.2.1.1.6.6 Medications Administered Section Constraints	39
	6.2.1.1.6.7 DICOM Object Catalog Section Constraints	40
115	6.2.1.1.7 Conformance.....	40
	6.2.2 Section Modules	41
	6.2.2.1 Medical History (Cardiac) Section	41
	6.2.2.1.1 Problem Concern Entry Constraints.....	41
	6.2.2.1.2 Simple Observation Entry Constraints.....	41
120	6.2.2.2 Family History (Cardiac) Section.....	42
	6.2.2.2.1 Simple Observation Entry Constraints	42
	6.2.2.3 Prior Results (Cardiac) Section	42
	6.2.2.3.1 Procedure Entry Constraints	43
	6.2.2.3.2 Simple Observation Entry Constraints	43
125	6.2.2.4 Cardiac Risk Assessment Section	43
	6.2.2.4.1 Simple Observation Entry Constraints.....	44
	6.2.2.5 Indications and Planned Procedure Section.....	44
	6.2.2.6 Cardiac Procedure Description Section	44
	6.2.2.6.1 Procedure Entry Constraints	45
130	6.2.2.6.2 Medication (medications administered) Constraints	45
	6.2.2.6.3 Medication (contrast/imaging agents) Constraints	45
	6.2.2.6.4 Simple Observation (procedure protocol) Constraints	46
	6.2.2.7 Procedure Findings (Cardiac) Section.....	47
	6.2.2.7.1 Simple Observation (ischemic heart disease) Constraints.....	48
135	6.2.2.7.2 Coronary Angiography Arterial Analysis Entry Constraints.....	49
	6.2.2.7.3 Simple Observation (LV function) Constraints.....	49
	6.2.2.7.4 Segmental Wall Analysis Entry Constraints	50
	6.2.2.7.5 Simple Observation (cardiac morphology) Constraints	50
	6.2.2.7.5.1 Pericardial Effusion Size Observation	53
140	6.2.2.7.5.2 Severity Observation.....	54
	6.2.2.7.5.3 Dissection targetSite Observation.....	54
	6.2.2.8 Assessment And Plan (Cardiac) Section	54
	6.2.2.8.1 Simple Observation (summary) Constraints	55
	6.2.2.9 Comparison to Prior Study (Cardiac) Section	55
145	6.2.2.9.1 Simple Observation (comparison) Constraints.....	55
	6.2.2.10 Key Images Section.....	56
	6.2.2.11 DICOM Object Catalog Section.....	56
	6.2.3 Header Content Modules	58
	6.2.3.1 Encompassing Encounter	58
150	6.2.3.1.1 Responsible Party.....	59
	6.2.3.1.2 Health Care Facility	60

	6.2.3.1.3 Referring Provider	60
	6.2.3.1.4 Physician of Record	61
	6.2.3.2 Order	62
155	6.2.3.3 Patient Demographics	62
	6.2.3.4 Service Event and Performer.....	63
	6.2.3.5 Legal Authenticator.....	65
	6.2.4 Entry Content Modules	66
	6.2.4.1 Diabetes Problem Entry	66
160	6.2.4.2 Diabetes Therapy Entry.....	66
	6.2.4.3 Angina Problem Entry.....	67
	6.2.4.3.1 Simple Observation (angina stability) Constraint	67
	6.2.4.3.2 Simple Observation (angina characteristics) Constraint.....	67
	6.2.4.4 Segmental Wall Analysis	68
165	6.2.4.4.1 Simple Observation (wall motion) Constraints	68
	6.2.4.4.2 Simple Observation (wall morphology) Constraints	69
	6.2.4.4.3 Simple Observation (viability) Constraints	69
	6.2.4.5 Coronary Angiography Arterial Analysis.....	69
	6.2.4.5.1 Simple Observation (stenosis) Constraints.....	70
170	6.2.4.5.2 Simple Observation (coronary findings) Constraints	71
	6.2.4.6 observationMedia Entry	71
	6.2.4.7 Planned Procedure Activity Entry	71
	6.3 VALUE SETS	73
	1.3.6.1.4.1.19376.1.4.1.5.1 Cardiac imaging services.....	73
175	1.3.6.1.4.1.19376.1.4.1.5.2 Cardiac imaging procedures	73
	1.3.6.1.4.1.19376.1.4.1.5.3 Referring Physician Specialty.....	73
	1.3.6.1.4.1.19376.1.4.1.5.4 Cardiac problems/concerns	74
	1.3.6.1.4.1.19376.1.4.1.5.5 Cardiac interventions.....	74
	1.3.6.1.4.1.19376.1.4.1.5.7 Angina Type	75
180	1.3.6.1.4.1.19376.1.4.1.5.10 Contrast Agents Classes for Adverse Reactions	75
	1.3.6.1.4.1.19376.1.4.1.5.11 Contrast Agents.....	75
	1.3.6.1.4.1.19376.1.4.1.5.13 Cardiac Lab Results.....	76
	1.3.6.1.4.1.19376.1.4.1.5.14 Cardiac Drug Classes.....	76
	1.3.6.1.4.1.19376.1.4.1.5.15 Drug Classes Used in Cardiac Procedure.....	77
185	1.3.6.1.4.1.19376.1.4.1.5.17 Rhythm Findings	77
	1.3.6.1.4.1.19376.1.4.1.5.18 ECG Findings	77
	1.3.6.1.4.1.19376.1.4.1.5.19 Myocardium Assessments	78
	1.3.6.1.4.1.19376.1.4.1.5.20 Wall motion.....	78
	1.3.6.1.4.1.19376.1.4.1.5.22 Cardiac Chamber Size Assessments.....	78
190	1.3.6.1.4.1.19376.1.4.1.5.23 Pulmonary Veins Assessments	79
	1.3.6.1.4.1.19376.1.4.1.5.24 Imaging Findings.....	79
	1.3.6.1.4.1.19376.1.4.1.5.25 Angina Characteristics	79
	1.3.6.1.4.1.19376.1.4.1.5.26 Cardiac risk scores	79
	1.3.6.1.4.1.19376.1.4.1.5.27 Cardiac Imaging Procedure Indications	80
195	1.3.6.1.4.1.19376.1.4.1.5.28 Cardiac Synchronization Techniques	81
	1.3.6.1.4.1.19376.1.4.1.5.29 Cardiac Shunt Types.....	81
	1.3.6.1.4.1.19376.1.4.1.5.30 ECG Summary Findings	81
	1.2.840.10008.6.1.48 DICOM CID 3015 - Coronary Arteries.....	81
	1.2.840.10008.6.1.57 DICOM CID 3261 - Stress Protocols	82
200	1.2.840.10008.6.1.211 DICOM CID 3710 - Coronary Dominance	83
	1.2.840.10008.6.1.218 DICOM CID 3717 - Myocardial Wall Segments	83
	1.2.840.10008.6.1.219 DICOM CID 3718 - Myocardial Wall Segments in Projection	84
	1.2.840.10008.6.1.223 DICOM CID 3722 - Diabetic Therapy	84
	1.2.840.10008.6.1.225 DICOM CID 3724 - Smoking History.....	85
205	1.2.840.10008.6.1.253 DICOM CID 3758 - Cardiovascular Family History	85
	1.2.840.10008.6.1.616 DICOM CID 12224 - Ultrasound Image Modes.....	86

	<i>1.2.840.10008.6.1.744</i>	<i>DICOM CID 3112 - Attenuation Correction</i>	86
	<i>1.2.840.10008.6.1.755</i>	<i>DICOM CID 3200 - Stress Test Procedure</i>	86
	<i>1.2.840.10008.6.1.759</i>	<i>DICOM CID 3204 - Stress Agents</i>	86
210	<i>1.2.840.10008.6.1.770</i>	<i>DICOM CID 3217 - Comparison Finding</i>	87
	<i>1.2.840.10008.6.1.781</i>	<i>DICOM CID 3235 - Perfusion Comparison Findings</i>	87
	<i>1.2.840.10008.6.1.782</i>	<i>DICOM CID 3236 - Tolerance Comparison Findings</i>	87
	<i>1.2.840.10008.6.1.783</i>	<i>DICOM CID 3237 - Wall Motion Comparison Findings</i>	88
	<i>2.16.840.1.113883.3.88.12.3221.5.2</i>	<i>HITSP Health Insurance Type</i>	88
215			

220 **Introduction**

This Supplement introduces a new Profile to the IHE Cardiology Technical Framework, with the overall design in Volume 1 and specific content in Volume 2. This Profile relies heavily on Content Profile concepts specified in the IHE Patient Care Coordination Technical Framework.

225 This content profile is motivated by cardiologists, who face an increasing demand from patient-care, data-quality and legislative perspectives to increase the usefulness and actionability of (discrete) clinical data across the various care-settings and stakeholders.

230 A solution for such interoperability is, however, not a simple undertaking. Unstructured textual data forms remains the predominate mechanism for information exchange among health care providers, and a good majority of data needed by physicians and other health care providers to make good clinical decisions is embedded in this free text. Efficient and effective interoperability therefore begins by identifying the most relevant clinical data.

Clinically-relevant cardiac imaging data is the key value proposition of this profile. The approach is to:

- 235
1. reuse the distribution and structuring work from the XDS (ITI domain), Medical summaries (PCC domain), and exchangeable procedure notes (HL7)
 2. extend it through adding and codifying the Key Data Elements for Cardiac Imaging¹ content identified by the American College of Cardiology (ACC) / American Heart Association (AHA) task-force on clinical data standards

240 The aim is to enable collection and distribution of the most clinically-relevant discrete data on the diagnostic imaging procedures common in cardiology. The usage of the discrete data is two-fold:

1. To enable individual test data to be more easily shared and used between care givers and systems
2. To enable population-based outcomes-based research on test effectiveness

245 These imaging procedures are used as key constituents of the patient’s treatment during cardiac encounters and disease management. Allowing a means to extract and exchange key cardiac measures across providers and their systems will be a huge advantage to providing a complete, accessible and actionable cardiac data set in front of cardiologists.

250 There are very successful quality-improvement programs in place by the professional bodies such as the ACC, AHA, and state registries concentrating on the most invasive, and expensive

¹ ACC/AHA/ACR/ASE/ASNC/HRS/NASCI/RSNA/SAIP/SCAI/SCCT/SCMR/SIR 2008 Key Data Elements and Definitions for Cardiac Imaging. JACC Vol. 53, No. 1, 2009.
(<http://content.onlinejacc.org/cgi/content/full/53/1/91>)

cardiac procedures. However, the vast majority of cardiology imaging usage is in the area of non-interventional diagnostic procedures. The aggregate volume and reimbursement level of these procedures mean that there must be significant automated collection/processing of key discrete data to make it economical to evaluate effectiveness of these procedures.

255 This supplement provides a framework to make progress on both these goals. This profile codifies extensive areas of diagnostic imaging indications, procedures, medications, observations, complications and findings and specifies how this discrete data can be exchanged to be used by both care-providers and automated data processing systems.

260 Future work is expected to extend the types of cardiac procedures that can be reported in a structured format to support existing population-based registries e.g., National Cardiovascular Data Registry (NCDR)-PCI for interventional procedures. Another key target of this future work includes electrophysiology, based on the forthcoming Electrophysiology (EP) key data elements recommendations from the Heart Rhythm Society (HRS).

Relationship to Workflow Profiles

265 Cardiac Imaging Report Content (CIRC) is a *content* profile – it is agnostic with respect to the workflow or data exchange mechanism in which the data is produced and handled.

270 Content Profiles define how the content used in a transaction is structured. The binding of the Content to an IHE transaction that is part of an IHE Workflow Profile specifies how this payload may influence the metadata or the behavior of the transaction. Content modules within the Content Profile then define the payloads. Content modules are transaction neutral, in that what they describe is independent of the transaction in which they are used, whereas content bindings explain how the payload influences the transaction metadata and/ or behavior.

275 The CIRC content is intended to be deployed, for example, in the Image Enabled Office (IEO) workflow profile for ambulatory care environments, the Displayable Reports (DRPT) workflow profile for in-patient environments, or the Cross-Enterprise Document Sharing (XDS) profile to propagate the content across organizational boundaries.

It is important to note that that key report-generation/distribution workflow aspects such as physician identification, insurance preauthorization, report routing and acknowledgement, and patient consent, are **out of scope** for this Content Profile.

280 Open Issues and Questions

#	Open Issue Description
11	A local extension <ihecard:accreditation> has been specified to allow physician or facility accreditation to be described. It there a more standard way to handle this?
12	LOINC code for a risk assessment document section is pending assignment.
13	Certain concepts represented by SNOMED compositions have been submitted for pre-coordinated codes.

Closed Issues

#	Closed Issue Description/ Resolution
1	<p>The Supplement uses the convention that a content module cannot have more than one parent (section 6.1.2), yet in section 6.2.1.1.2 the Cardiac Imaging Report Template is described as being a specialization (descendant) of both the Medical Document template and the Procedure Note template. This inconsistency needs to be resolved prior to this profile being published for Trial Implementation.</p> <p><i>Answer: The restriction of having no more than a single parent has been removed from the Trial Implementation version.</i></p>
2	<p>There are a number of instances where codes are missing (or unaccounted for) in some of the value sets listed in the 6.3.x subsections. These gaps need to somehow be resolved prior to this profile being published for Trial Implementation. Examples of missing codes include:</p> <ul style="list-style-type: none"> a. Missing code for pharma echo Stress tests (just exercise stress) b. Missing codes for some Echo and Cardiac MR procedure acquisition types c. Missing codes for a couple of contrast agents d. Missing codes for LVEF Assessment concepts e. Missing codes for Cardiac Risk Assessment scores (e.g., Framingham risk assessment score). List of concepts has been submitted for inclusion in LOINC. <p><i>Answer: Missing and unaccounted for codes have been assigned/ filled in in the necessary code set definitions.</i></p>
3	<p>How should non-DICOM "images" (such as Bull's eye diagrams and coronary tree diagrams) be included in the Cardiac Imaging Report?</p> <p><i>Answer: The supplement describes a method to inline encode non-DICOM images such as coronary trees or "bull's eye diagrams" using the CDA 'observationMedia' entry (see section 6.2.4.6).</i></p>
4	<p>There are currently no specific requirements for Content Consumers with respect to having to handle multiple coding schemes. Should there be?</p> <p><i>Answer: The use of specific coding schemes to codify the wide variety of concept included in the CIRC Content Profile is frequently a national, regional or even local decision based on regulation or legislation. Thus the Cardiology Technical Committee documented the most common, internationally applicable codes in the value sets defined, but stopped short of mandating specific coding scheme in this specification.</i></p>

#	Closed Issue Description/ Resolution
5	<p>The current profile supplement does not describe how to identify profile conformance, especially as the template evolves. Is this something that should be covered in this Content Profile, and if so, how should that be done?</p> <p><i>Answer: The profile will not describe an explicit method for conformance claims to specific templates or template elements. Rather it follows the existing convention used by other CDA implementation guides that substantive changes in requirements to a template at any level necessitates the creation of a new OID to identify the changed template as well as a new OID for any other template that invokes the changed template as a required element. (Changes to optional elements generally do not affect conformance.)</i></p>
6	<p>The supplement currently does not describe how to reference the prior study referenced in the observation(s) within Comparison section. Assuming this referencing should be required if a comparison section exists, should that reference be done based on some DICOM identifier (e.g., study UID), or some human identifiable item (e.g., study date)?</p> <p><i>Answer: The comparison study template has been updated to require that the narrative description of the comparison study include the exam date and the exam type in order to provide "human identifiable" information in the report.</i></p>
7	<p>Does this supplement need to describe how the content data gets mapped into XDS meta-data (to see a similar situation, refer to medical summary)? Examples are format codes/code-lists that differentiate medical summary from cardiac report. It is believed that this is "covered" by the reference made to the PCC TF specification for XDS metadata mapping (specific reference is to XDS-MS). The Cardiology Technical Committee solicits feedback from the PCC Technical Committee on this approach.</p> <p><i>Answer: No negative feedback was received during public comment, so no changes were made to the document.</i></p>
8	<p>There are some missing details for describing Cardiac Findings in section 6.2.2.11. These gaps in detailed data modeling need to be resolved prior to this profile being published for Trial Implementation. Such missing information is highlighted in yellow.</p> <p><i>Answer: All known gaps in the findings section template have been filled.</i></p>

#	Closed Issue Description/ Resolution
9	<p>The supplement does not aim to be comprehensive for all downstream uses of discrete data - from next-level analysis, discrete data warehousing of clinical record, to registry submission.</p> <p><i>Answer: The Cardiology Technical Committee believes this is acceptable for the following reasons:</i></p> <ul style="list-style-type: none"> <i>a. The intent of this content profile is to capture structure and codify the key discrete data that is today conveyed in cardiology imaging reports.</i> <i>b. There is no widely accepted standard for report composition across the broad range of imaging procedures in cardiology.</i> <i>c. Furthermore, registry-submission and clinical records have broader data scope than individual cardiac procedure reports. The work here is an initial attempt to support the communication of discrete cardiac data throughout and between enterprises, it will require extension and extensive collaboration in future updates to the IHE-TF.</i>
10	<p>Should this Cardiac Imaging Report Template inherit its definition from both the Procedure Report Implementation Guide and the Basic Diagnostic Imaging Report Implementation Guide?</p> <p><i>Answer: This document is not a specialization of the HL7 Basic Diagnostic Imaging Report template due to conflicts with two Procedure Note requirements (format of serviceEvent/effectiveTime, and title on DICOM Catalogue section). When and if these are resolved, an instance may also comply to the Diagnostic Imaging Report.</i></p>

Volume 1 – Integration Profiles

285

Add the following bullet to the end of the bullet list in section 1.7

1.7 Current Year Scope

- Added the Cardiac Imaging Report Content which describes the content and format of the clinical report created for a cardiac imaging procedure.

290

Add the following to Table 2-1. Update table number to 2.1-1

2.1 Dependencies among Integration Profiles

Table 2.1-1. Cardiology Integration Profiles Dependencies

Integration Profile	Depends on	Dependency Type	Comments
...			
Cardiac Imaging Report Content	Consistent Time	Content Creator is required to be grouped with Time Client actor	

295

Add the following section to section 2.2

2.2.11 Cardiac Imaging Report Content (CIRC)

The Cardiac Imaging Report Content (CIRC) Profile specifies the content structure for a clinical report of a cardiology imaging exam, recorded in a DICOM Study. Such exams include:

- Echocardiography (transthoracic - TTE, transesophageal - TEE, and TTE stress)
- Cardiac computed tomography (angiography - CCTA, and coronary artery calcium scoring - CACS)
- Cardiac magnetic resonance (angiography - MRA, and MR stress)
- Cardiovascular nuclear medicine (SPECT myocardial perfusion, positron emission tomography - PET)
- Diagnostic coronary catheter based fluoroscopy (interventional coronary angiography - ICA, and left ventriculography - LVG)

The CIRC Profile specifies the use of an HL7 Clinical Document Architecture (CDA) format for the report.

310

Not included in the scope of this profile are non-imaging studies (e.g., ECG), electrophysiology procedures, and non-cardiology procedures (e.g., peripheral angiography). Such use cases may be supported by other similar content profiles.

315 *Add the following section to section 2.3*

2.3 Actor Descriptions

...

Content Creator – a system that creates documents in accordance with a content profile

Content Consumer – a system that receives documents in accordance with a content profile

320 ...

Table 2.3-1. Integration Profile Actors

Actor \ Integration Profile	...	<u>CIRC</u>
...		
<u>Content Creator</u>		<u>X</u>
<u>Content Consumer</u>		<u>X</u>

Add Section 11

11 Cardiac Imaging Report Content (CIRC) Integration Profile

325 The Cardiac Imaging Report Content (CIRC) Profile specifies the content structure for a clinical report of a cardiology imaging exam, recorded in a DICOM Study. Such exams include:

- Echocardiography (transthoracic - TTE, transesophageal - TEE, and TTE stress)
- Cardiac computed tomography (angiography - CCTA, and coronary artery calcium scoring - CACS)
- 330 • Cardiac magnetic resonance (angiography - MRA, and MR stress)
- Cardiovascular nuclear medicine (SPECT myocardial perfusion, and positron emission tomography - PET)
- Diagnostic coronary catheter based fluoroscopy (interventional coronary angiography - ICA, and left ventriculography - LVG)

335 The CIRC Profile specifies the use of an HL7 Clinical Document Architecture (CDA) format for the report. This format supports both the human readable narrative historically used for clinical reports, as well as a substantial set of discrete data elements that may be used for longitudinal or population analysis or other computer processing.

340 **Note:** It is expected that future evolution of cardiology reporting will incorporate more robust and extensive sets of data elements. This Profile is a first step beyond simple narrative to interoperable discrete data elements.

Not included in the scope of this profile are non-imaging studies (e.g., ECG), electrophysiology procedures, interventional therapeutic procedures (e.g., angioplasty), and non-cardiology procedures (e.g., peripheral angiography). Such use cases may be supported by other similar content profiles.

345 It is assumed that there is a DICOM Study associated with the exam. If there is not a DICOM Study, this report content may not be appropriate. In addition to reference images, the DICOM Study data may include discrete data elements encoded in DICOM Structured Report information objects that may be transcoded into the discrete data elements specified in this Profile. (See the Evidence Documents Profile and its cardiology options in Section 7.)

350 The CIRC Profile does not presume to describe the complete content of an imaging study report. It does provide the framework of high level section titles and a set of discrete data elements. Within that framework reports can be created with the clinical content desired by their authors, including additional discrete data elements. In general, there are no constraints on the narrative text and figures that the cardiologist could include in the report document, although there are
355 requirements on minimum data elements reflecting expert consensus (ACC/AHA/et al. 2008 Key Data Elements for Cardiac Imaging).

This profile also does not provide all of the details necessary to construct a CDA compliant document. Please refer to the HL7 CDA Release 2 Standard.

11.1 Actors/ Transactions

360 There are two actors in this profile, the Content Creator and the Content Consumer. Content is created by a Content Creator and is to be consumed by a Content Consumer. The sharing or transmission of content from one actor to the other is addressed by the appropriate use of other IHE profiles, and is out of scope of this profile; hence there is no transaction per se defined for this profile.



365

Figure 11.1-1 Cardiac Imaging Report Template Actor Diagram

Note: The primary intended transmission mechanism in the intra-institutional context is the IHE Displayable Reports Profile (DRPT), and in the inter-institutional context the IHE Cross Enterprise Document Sharing Profiles (XDS, XDM and XDR). A Report Creator, Document Source or a Portable Media Creator of those profiles may embody the Content Creator Actor. A Document Consumer, a Document Recipient or a Portable Media Importer may embody the Content Consumer Actor.

370

11.1.1 Actor Groupings

375 The Content Creator shall be grouped with Time Client actor of the IHE IT Infrastructure Consistent Time Profile, as specified in ITI TF-1:7. This allows the Legal Authentication timestamp to be accurate.

11.2 Cardiac Imaging Report Content Profile Options

380 Options that may be selected for this Content Profile are listed in Table 11.2-1 along with the Actors to which they apply. Dependencies between options when applicable are specified in notes.

Table 11.2-1 CIRC Profile Options

Actor	Option Name	Optionality	Section
Content Consumer	View Option	O (see 11.2.1)	PCC TF-2 :3.1.1
	Document Import Option	O (see 11.2.1)	PCC TF-2 :3.1.2
	Section Import Option	O (see 11.2.1)	PCC TF-2 :3.1.3
	Discrete Data Import Option	O (see 11.2.1)	PCC TF-2 :3.1.4
Content Creator	<i>No options defined</i>		

11.2.1 Content Consumer Options

385 The Content Consumer actor is required to support at least one of the View or Discrete Data Import options. The Document Import and Section Import options, if implemented, also require the View option. These options as specified in the PCC Technical Framework assume use of XDS or related profiles for transport; this Profile specifies bindings to other workflow profiles (see Section 11.5), and these options should be interpreted as applicable with any binding. See
390 also Section 11.6.2.

11.3 Use Cases

11.3.1 Compile and transfer report of cardiac imaging procedure with use of key imaging data measures

395 This use case addresses the generation and transfer of a cardiac imaging report based on the ACC/AHA Key data Elements for Cardiac Imaging¹. The initial content, structure and coding of the report to support this use case are detailed as part of this profile (see IHE CARD TF-2: 6 Content Modules). However various reporting system implementations, institute reporting guidelines and individual Reporting Physician usage may result in some variability in the specific report content provided.

400 **Pre conditions**

The systems underlying the data collection and management for the various elements of the procedure report has all the *mandatory* data elements identified using codes, and are expected to be the source for the information used in creating the *majority* of the structured imaging report document.

405 **Main Flow**

- Cardiologist reviews and/or records the codified
 - procedures and protocols used in the procedure
 - Image and related data generated from the various modalities and monitoring equipment used during the procedure so that they key physiological measures, acquired and derived are present in line with the ACC/AHA guidelines.
 - Other relevant patient characteristics
 - Medications documented for the patient both pre and during procedure.
 - Indications and observations/complications noticed during the procedure.
 - Findings, assessment and plan
- 415 • Cardiologist approves the procedure report and this marks it ready for distribution
- The content creator system will format the report appropriately (this profile) and send it via one of the IHE mechanisms to a content consumer system (an appropriate workflow profile).

Post conditions

420 The subsequent clinical stakeholder (system) receives the Document for import, processing and optionally viewing of the data.

11.3.2 Perform discrete data-analysis on procedure report content

425 The goal of this use case is to assist data collection for comparative and research purposes. Based on a report generated in the previous use cases an advanced medical data analysis system collects discrete data from multiple patients and their procedure, e.g., for cardiac Clinical Decision Support or for advance lifetime patient records.

Pre conditions

The content consumer (e.g., an advanced medical data analysis system) received a Cardiac Imaging Report with coded/structured content as defined in IHE CARD TF-2: 6.2 CDA Release 2 Content Modules.

430 **Main Flow**

The consuming system collects and processes the data from the various reports it receives and extracts those relevant data for either:

- A specific clinical concern for a population e.g., pre-populating a procedure-specific registry; extracting a data subset for a specific research question.
- 435 • A more comprehensive longitudinal patient record (e.g., an EMR) which can provide trending over time on an individual patient’s key cardiac measures.

Post conditions

440 The content consumer generated new (derived) data for use by others. The type of data generated is out of the scope of this *profile*.

11.3.3 Review Procedure Report

A secondary use-case addressed by this profile involves the direct human use of the procedure report. In most practical cases this will be:

- 445 • The referring physician who instigated/ordered the procedure, and other healthcare providers who manage subsequent patient care activities
- Another person involved in downstream clinical or administrative data processing e.g., someone validating/source-checking for QA the original report as part of JCAHO audits, or pre-submission checking on the original reporting data against the case-data imported in the NCDR-PCI registry-submission application

450 Pre conditions

- The Reviewing Physician consumer has a system (EMR or other) capable of importing and displaying the received report in a clinically useful format
- The Cardiac Imaging Report has been received at this system

Note: This profile does not assume any explicitly specified relationship between the creator and consumer.

455 Main Flow

- The reviewing physician selects the report of his patient and opens it for review
- The system displays the human readable content for the reviewing physician to review

Post conditions

The Reviewer has extracted (visually) the necessary information from the report.

460 11.4 Security Considerations

Security considerations are dealt with by the transport mechanism (e.g., XDS, DRPT) and are outside the scope of this content profile. See PCC TF-1: 3.8

11.5 Grouping

465 Content modules describe the content of a payload found in an IHE transaction. Content profiles are transaction neutral. They do not have dependencies upon the transaction that they appear in.

A Content Creator or Content Consumer must be grouped with appropriate actors from workflow profiles that manage interchange of clinical data. Such groupings are described in this section.

470 Content profiles may impose additional requirements on the transactions used when grouped with actors from other IHE Profiles. The metadata sent in the document sharing or interchange messages has specific relationships to the content of the clinical document described in the

content profile. These mappings between the workflow metadata and the content attributes are described in Volume 2 of the Technical Framework.

11.5.1 Content Bindings for Displayable Reports (DRPT) and Image Enabled Office (IEO) Profiles

475 CDA documents using the CIRC content may be exchanged between a Report Creator and a Report Manager, as defined in the Displayable Reports (DRPT) and Image Enabled Office (IEO) Profiles using the Encapsulated Report Submission [CARD-7] transaction. In these cases, the CIRC Content Creator actor is grouped with the DRPT or IEO Report Creator actor, and the CIRC Content Consumer actor is grouped with the DRPT or IEO Report Manager actor.

11.5.2 Content Bindings for XDS, XDM, XDR, and XDS-I

480 It is expected that the transfers of care will occur in an environment where the physician offices and hospitals have a coordinated infrastructure that serves the information sharing needs of this community of care. Several mechanisms are supported by IHE profiles:

- 485 • A registry/repository-based infrastructure is defined by the IHE Cross Enterprise Document Sharing (XDS) and other IHE Integration Profiles such as patient identification (PIX & PDQ) and notification of availability of documents (NAV). An extension for imaging study exchange is Cross Enterprise Document Sharing for Imaging (XDS-I).
- A media-based infrastructure is defined by the IHE Cross Enterprise Document Media Interchange (XDM) profile.
- 490 • A reliable messaging-based infrastructure is defined by the IHE Cross Enterprise Document Reliable Interchange (XDR) profile.
- All of these infrastructures support Security and privacy through the use of the Consistent Time (CT) and Audit Trail and Node Authentication (ATNA) profiles.

495 For more details on these profiles, see the IHE IT Infrastructure Technical Framework, and the IHE Radiology Technical Framework for XDS-I.

Document Source and Document Consumer Actors from the ITI XDS, XDM and XDR profiles are logically grouped with the CIRC Content Creator and Content Consumer actors, respectively.

11.5.3 Content Binding for Portable Data for Imaging (PDI)

500 CDA documents using the CIRC content may be exchanged on interchange media in accordance with the Portable Data for Imaging (PDI) profile. Such documents may be encapsulated within DICOM SOP Instances, or may be native CDA documents, as described in the IHE Radiology Technical Framework. In this case, the CIRC Content Creator actor is grouped with the PDI Portable Media Creator actor, and the CIRC Content Consumer actor is grouped with the PDI Display or Portable Media Importer actors.

11.5.4 Content Binding for Retrieve Form for Data Capture (RFD)

505 A CDA document may be used for pre-population of a data entry form managed by actors of the Retrieve Form for Data Capture (RFD) Profile. In particular, the CIRC content, as a carrier of

discrete encoded data, may be used to prepopulate data entry forms for cardiovascular data registries. The CIRC profile has been developed with key data elements that support common research related data fields. This profile, however, does not provide mapping between CIRC field content and any specific registry field content. In this case, the CIRC Content Consumer actor is grouped with the RFD Form Manager actor for the purpose of extracting discrete data from the report to pre-populate the data capture form.

11.5.5 Relationship to Document Digital Signature (DSG)

When a Content Creator Actor needs to digitally sign a document in a submission set, it may support the Digital Signature (DSG) Content Profile as a Document Source. When a Content Consumer Actor needs to verify a Digital Signature, it may retrieve the digital signature document and may perform the verification against the signed document content.

11.6 Requirements of CIRC Actors

This section describes the specific requirements for each Actor defined within this profile. Specific details can be found in Volume 1 and Volume 2 of the technical framework.

Note: The IHE PCC Technical Framework includes additional requirements for actors associated with its Content Profiles. Those profiles are generally intended to be used in cross-enterprise exchanges of data, and therefore include actor requirements related to that environment. The CIRC profile has a significant intra-institutional intended use, and hence requirements not necessary for that environment have been excluded.

11.6.1 Content Creator

1. A Content Creator shall be able to create a Cardiac Imaging Report Document according to the specifications for that content profile found in CARD TF-2.
2. A Content Creator shall be grouped with the Time Client Actor, and shall synchronize its clock with a Time Server.

11.6.2 Content Consumer

1. A Content Consumer shall be able to consume (receive and process) a Cardiac Imaging Report document.
2. A Content Consumer shall implement the View Option or Discrete Data Import option, or both.
3. A Content Consumer that implements the Document Import or Section Import Option shall implement the View Option as well.
4. A Content Consumer that implements the View option shall be able to:
 - a. Demonstrate rendering of the document for display.
 - b. Print the document.
 - c. Display the document with its original style sheet.
 - d. Support traversal of any links contained within the document.

5. A Content Consumer that implements the Document Import Option shall:

a. Store the document.

545

b. Demonstrate the ability to access the document again from that storage.

6. A Content Consumer that implements the Section Import Option shall offer a means to import one or more document sections into the patient record as free text.

7. A Content Consumer that implements the Discrete Data Import Option shall offer a means to import structured data from one or more sections of the document.

550

Volume 2 – Transactions and Content Modules

Update section 3

3 Framework Overview

555 The IHE Technical Framework is based on actors that interact through transactions; **those transactions may be further qualified with respect to their content.**

3.1 Actors and Transactions

Actors are information systems or components of information systems that produce, manage, or act on information associated with operational activities in the enterprise.

560 Transactions are interactions between actors that transfer the required information through standards-based messages.

Specific sets of actors and transactions are specified in the Integration Profiles (see CARD TF-1). **While transactions are described in this volume (CARD TF-2), the role and implementation of these transactions require the understanding of the Integration profile they support.**

565

Add new section 3.2

3.2 Content Modules

570 There is often a very clear distinction between the transactions in a messaging framework used to package and transmit information, and the information content actually transmitted in those messages. This is especially true when the messaging framework begins to move towards mainstream computing infrastructures being adopted by the healthcare industry.

In these cases, the same transactions may be used to support a wide variety of use cases in healthcare, and so more and more the content and use of the message also needs to be profiled, sometimes separately from the transaction itself. Towards this end IHE has developed the concept of a Content Integration Profile.

575

Content Integration Profiles specify how the payload of a transaction fits into a specific use of that transaction. A content integration profile has three main parts. The first part describes the use case (this is found in Volume 1 in the definition of each Profile). The second part is a Content Module (found in this Volume 2), which describes the payload of the transaction; a content module is specified so as to be independent of the transaction in which it appears. The third part is binding to a specific IHE transaction, which describes how the content affects the

580

transaction. The binding of CDA-based medical documents to workflow transactions is described in the Profile definition in Volume 1 (e.g., see IHE CARD TF-1:11.5).

585

Add new section 5

5 Namespaces and Vocabularies

This section lists the namespaces and identifiers defined or referenced by the IHE Cardiology Technical Framework, and the vocabularies defined or referenced herein.

590

codeSystem	codeSystemName	Description
1.3.6.1.4.1.19376.1.5.3.1	IHE PCC Template Identifiers	This is the root OID for all IHE PCC Templates. A list of PCC templates can be found in IHE PCC TF-2:6.2 (CDA Release 2.0 Content Modules).
1.3.6.1.4.1.19376.1.4.1	IHE Cardiology Template Identifiers	This is the root OID for all IHE Cardiology Templates.
2.16.840.1.113883.5.112	RouteOfAdministration	The HL7 RouteOfAdministration coding scheme
2.16.840.1.113883.5.7	ActPriority	The HL7 ActPriority coding scheme
2.16.840.1.113883.6.1	LOINC	Logical Observation Identifier Names and Codes
2.16.840.1.113883.6.96	SNOMED CT	SNOMED Controlled Terminology
2.16.840.1.113883.6.88	RxNorm	RxNorm
1.2.840.10008.2.16.4	DCM	DICOM Controlled Terminology; PS 3.16 Content Mapping Resource, Annex D
2.16.840.1.113883.6.24	MDC	ISO/IEEE 11073 Medical Device Nomenclature
2.16.840.1.113883.3.26.1.5	NDF-RT	National Drug File Reference Terminology (NCI version)
2.16.840.1.113883.11.19465	nuccProviderCodes	National Uniform Codes Council Healthcare Provider Terminology
2.16.840.1.113883.6.255.1336	X12DE1336	Insurance Type Code (ASC X12 Data Element 1336)
2.16.840.1.113883.6.256	RadLex	RadLex (Radiological Society of North America)

5.1 IHE Format Codes

The table below lists the format codes, root template identifiers and media types used by the IHE Profiles specified in the Cardiology Technical Framework.

595

Note that the code system for these codes is **1.3.6.1.4.1.19376.1.2.3** as assigned by the ITI Domain for codes used for the purposes of cross-enterprise document sharing (XDS).

Profile	Format Code	Media Type	Template ID
---------	-------------	------------	-------------

IHE Cardiology Technical Framework Supplement – Cardiac Imaging Report Content (CIRC)

Profile	Format Code	Media Type	Template ID
Cardiac Imaging Report	urn:ihe:card:imaging:2011	text/xml	1.3.6.1.4.1.19376.1.4.1.1.1

600 *Add new section 6*

6 Content Modules

6.1 Conventions

6.1.1 Content Module Conventions

6.1.1.1 Cardinality Constraints

605 Within Section 6, the following conventions are used to describe data element cardinality constraints.

The cardinality expresses the number of times an attribute or association may appear in a CDA document instance that conforms to the specifications described within section 6. Cardinality is expressed as a minimum and a maximum value separated by '.', and enclosed in '[]', e.g.,

610 '[0..1]'.

Minimum cardinality is expressed as an integer that is equal to or greater than zero. If the minimum cardinality is zero, the element need only appear in message instances when the sending application has data with which to value the element. Mandatory elements must have a minimum cardinality greater than zero.

615 The maximum cardinality is expressed either as a positive integer (greater than zero and greater than or equal to the minimum cardinality) or as unlimited using an asterisk ("*").

6.1.1.2 Data Element Optionality Constraints

620 Within Section 6, the following conventions are used to describe data element optionality constraints. Where applicable, the "interaction" between cardinality constraints and optionality constraints are also described below.

Table 6.1.1.2-1 Data Element Optionality Constraints

Optionality	Description
M	A "Mandatory" section, entry or data element is one that SHALL always be provided. If there is information available, the element must be present and non-null. If there is no information available, or it cannot be transmitted, the data element must contain a value indicating the reason for omission of the data. Note that any element declared to be "Mandatory" must also be "Required" and have a minimum cardinality of one.
R	A "Required" section, entry or element SHALL be included in the document if its minimum cardinality is one. If the data exists, the sending application SHALL send it as a non-null value or a non-empty element. If the data does not exist and if the minimum cardinality is greater than zero, then the sending application SHALL send an appropriate null value. Only if data does not exist for a required element and that element has a minimum cardinality of 0 MAY the required element be

Optionality	Description
	<p>omitted in a document.</p> <p>In all cases, if a required element is present in a document received by an actor claiming support for the Profile, then it SHALL be correctly processed by the receiving actor. A receiving actor SHALL NOT raise an error due to the absence of a required element with a cardinality of 0, although it MAY issue a warning that required information is missing.</p> <p>For required elements, conforming applications must demonstrate their ability to provide and communicate not null values. Receiving applications must demonstrate their ability to receive and process (e.g., store, or display to users) not null values for required elements.</p> <p>This is equivalent to a SHOULD requirement.</p>
O	<p>An optional data element is one that MAY be provided, whether the information is available or not. If the implementation elects to support this optional section, then its support shall meet the requirement set forth for the "Required" or R.</p>
C	<p>A conditional data element is one that is required, or optional, depending upon other conditions. These will have further notes explaining when the data element is required.</p>

Note: The definitions of M, R, and O are consistent with HL7 v3 Conformance profiles, but differ slightly from the 2010 and earlier versions of IHE Patient Care Coordination Content or Workflow profiles. It is expected that all IHE Technical Framework documents will converge to these HL7-based definitions.

625

6.1.1.3 Coded Terminology Values

Coded terminology values are used extensively, and are encoded in CDA documents using the CD (Concept Descriptor) data type. Generally, these values are specified in Profile requirements using a triplet of the code value (encoded in XML attribute `code`), the coding scheme (encoded in XML attribute `codeSystemName`), and the code meaning (encoded in XML attribute `displayName`). When necessary to disambiguate such a triplet from the rest of the specification text, it may be enclosed in curly braces, e.g., {160245001, SNOMED CT, “No current problems or disability”}.

Representation of a coded terminology value in the CD data type requires encoding of the coding scheme OID in XML attribute `codeSystem`. For readability, these OIDs are not elaborated in the specification text. Content Creator actors must use the appropriate OIDs from Section 5 in encoding CD data type values.

Unless otherwise specified, value sets are specified with STATIC stability and have CWE (Coded With Extensibility) coding strength, as defined in the HL7 Core Principles and Properties of v3 Models. That is, the version of the value set as of the date of publication of the Profile is binding, and an implementation may use coded concepts not present in the value set.

6.1.2 Structure of Content Modules

For CDA Release 2 the Content Modules are organized by document, section, entry, and header elements.

645




Figure 6.1.2-1 CDA R2 R-MIM with location of Document, Sections, and Entries

Each content module is defined in terms of constraints that must be obeyed by instances of that content module, in effect a contract between the Content Creator and the Content Consumer. Each content module has a name, also known as its template identifier. The template identifiers are used to identify the contract implied by the content module.

Content modules may inherit features of other content modules of the same type (Document, Section, or Entry) by defining the parent content module that they inherit from. They may not inherit features from a different type. Although information in the CDA Header is in a different location than information in a CDA Entry, these two content modules are considered to be of the same type, and so may inherit from each other when necessary.

Each content module has a list of data elements that are mandatory (M), required if known (R), optional (O), and conditional (C). The presentation of this information varies with the type of content module, and is described in more detail below. Additional data elements may be provided by the sender that are not defined by a specific content module, but the receiver is not required to interpret them. Thus, it is not an error to include more than is asked for, but it is an error to reject a content module because it contains more than is defined by the template. This allows values to be added to the content modules delivered in this framework, through extensions to it that are not defined or profiled by IHE. It further allows content modules to be defined later by IHE that are refinements or extensions over previous content modules.

In order to retain this capability, constraints that apply to any content module will always apply to any content modules that inherit from it. Thus, the "contracts" are always valid down the

670 inheritance hierarchy. Second, data elements of a content module will rarely be deprecated. This
will usually occur only in the cases where they have been deprecated by the base standard. While
any specific content module has a limited scope and set of use cases, deprecating the data
675 element prevents any future content module from taking advantage of what has already been
defined when a particular data element has been deprecated simply because it was not necessary
in the original use case.

675 **6.1.2.1 Document Content Modules**

Each **document** content module will define the appropriate codes used to classify the document,
and will also describe the specific section and header data elements that are included. The code
used to classify it is specified using an external vocabulary, typically LOINC in the case of CDA
680 Release 2 documents. The set of data elements that make up the document are defined, including
the whether these data elements must, should or may be included in the document. Each data
element is mapped to a lower level content module via a template identifier, and the document
content module will further indicate whether these data elements are mandatory, required if
known or optional. Thus, a document content module contains as constraints:

- The template identifier of the parent content module when there is one.
- 685 • The LOINC code or codes that are used to classify the document.
- A possibly empty set of mandatory, required if known, and optional header content modules,
and their template identifiers.
- A possibly empty set of mandatory, required if known, and optional section content modules,
and their template identifiers.
- 690 • Other constraints as necessary.

The order of section content modules is not specified; sections may appear in any order, and may
be nested, in accordance with local implementation style specifications.

6.1.2.1.1 Document Content Module Table

The Document Content Module is specified using the following table.

695

Template ID				
Parent Template				
General Description				
Document Code				
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Header Elements				
Sections				

This table implies the following conformance statements:

1. The document SHALL include the specified Template ID in the <templateID> element of the <clinicalDocument> act element (the CDA root act).
- 700 2. The document SHALL conform to all the requirements of the specified Parent Template(s).
3. The document SHALL include the specified Document Code in the <code> element of the <clinicalDocument> act element, except if the specified Document Code includes the keyword “SHOULD or “MAY”; in the latter case, this requirement is relaxed to the requirement strength of those keywords.
- 705 4. The document SHALL include the specified Header Elements in accordance with their specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1), in accordance with the specified Template ID and further constraints specified in the identified Technical Framework section.
- 710 5. The document SHALL include the specified Sections in accordance with their specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1), in accordance with the specified Template ID and further constraints specified in the identified Technical Framework section.

Note: The further constraints are typically specific value sets to be applied to code elements in the template.

715 The Document Content Module table may be supplemented with additional specific conformance requirements.

6.1.2.2 Section Content Modules

720 **Section** content modules will define the content of a section of a clinical document. Sections will usually contain narrative text, and so this definition will often describe the information present in the narrative, although sections may be wholly comprised of subsections.

Sections may contain various subsections, and these may be mandatory, required if known or optional. Sections may also contain various entries, and again, these may be mandatory, required if known, or optional. A section may not contain just entries; it must have at least some narrative text or subsections to be considered to be valid content.

725 Sections can inherit constraints from another parent section content module. Sections are classified using an external vocabulary (again typically this would be LOINC, although in some cases DICOM), and so the list of possible section codes is also specified. Sections that inherit from another section module will specify the same section code(s) as its parent, unless it further restricts the type of section to smaller set of codes.

730 Thus, a section content module will contain as constraints:

- The template identifier of the parent content module when there is one.
- The code or codes that shall be used to classify the section.
- A possibly empty set of mandatory, required if known, and optional section content modules, and their template identifiers for the subsections of this section.
- 735 • A possibly empty set of mandatory, required if known, and optional entry content modules, and their template identifiers.
- Other constraints as necessary.

6.1.2.2.1 Section Content Module Table

The Section Content Module is specified using the following table.

740

Template ID				
Parent Template				
General Description				
Section Code				
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Subsections				
Entries				

This table implies the following conformance statements:

1. The section SHALL include the specified Template ID in the <templateID> element of the <section> act element.
- 745 2. The section SHALL conform to all the requirements of the specified Parent Template.

3. The section SHALL include the specified Section Code in the <code> element of the <section> act element, except if the specified Section Code includes the keyword “SHOULD or “MAY”; in the latter case, this requirement is relaxed to the requirement strength of those keywords.
- 750 4. The section SHALL include the specified Subsections in accordance with their specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1), in accordance with the specified Template ID and further constraints specified in the identified Technical Framework section.
- 755 5. The section SHALL include the specified Entries in accordance with their specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1), in accordance with the specified Template ID and further constraints specified in the identified Technical Framework section.

The Section Content Module table may be supplemented with additional specific conformance requirements.

760 **6.1.2.2.2 Observation Entry Constraint Table**

Constraints on Entries may be further specified using the following table. The template for the entry (typically the IHE PCC Simple Observation template) is specified by the invoking table, for which this table provides additional constraint specifications. Multiple rows may be present in the table to specify constraints on multiple entries based on a template invoked with
765 cardinality greater than 1.

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set

This table implies the following conformance statements:

1. There SHALL be entries in accordance with each row in the table in accordance with the specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1).
- 770 2. Conditional (C) entries SHALL be present in accordance with the specified Exam Type Condition.

Note: The exam type is specified in the CDA Header in the documentationOf / serviceEvent / code element.

3. The entry SHALL include the specified observation / code element value, and the specified targetSiteCode and/or methodCode elements if present in the table.

775 Note: The codes may be specified as a value selected from an identified Value Set.

4. The entry SHALL include a value of the specified Data Type.
5. If Data Type is PQ, the entry value SHALL use the specified Unit of Measure.
6. If Data Type is CD, the entry value SHALL be selected from the specified Value Set.

Notes: 1. The code may be specified as a single value, rather than as a selection from a Value Set.

780 2. The Value Set table entry may indicate the presence of additional constraints, e.g., for specification of severity, by a '+' and a constraint type. Such additional constraints will have specific requirements specified outside the table.

6.1.2.3 Entry and Header Content Modules

785 **Entry** and **Header** content modules are the lowest level of content for which content modules are defined. These content modules are associated with classes from the HL7 Reference Information Model (RIM). These "RIM" content modules will constrain a single RIM class. Entry content modules typically constrain an "Act" class or one of its subtypes, while header content modules will normally constrain "Participation", "Role" or "Entity" classes, but may also
790 constrain an "Act" class.

Entry and Header content modules describe the mandatory, required if known, and optional XML elements and attributes that are present in the CDA Release 2 instance. Header and Entry content modules may also be built up using other Header and Entry content modules. An entry or header content module may also specify constraints on the vocabularies used for codes found in
795 the entry, or data types for the values found in the entry. Thus, an entry or header content module will contain as constraints:

- The template identifier of the parent content module when there is one.
- A description of the XML elements and attributes used in the entry, along with explanations of their meaning.
- 800 • An indication of those XML elements or attributes that are mandatory, required if known, or optional.
- Vocabulary domains to use when coding the entry.
- Data types used to specify the value of the entry.
- Other constraints as necessary.

805 **6.1.2.3.1 Header Content Module Table**

A Header Content Module is specified using the following table.

Template ID					
Parent Template					
General Description					
Header Element					
Code					
Opt	Participation	Description	Template	Spec Document	Con-straint

This table implies the following conformance statements:

1. The specified Header Element SHALL be present in the CDA header.

810 Note: This is limited by the Cardinality and Optionality of the header data element as specified in the template that invokes this Content Module.

2. The header data element SHALL include the specified Template ID in the <templateID> element of the relevant act element.

815 3. The header data element SHALL conform to all the requirements of the specified Parent Template.

4. The header data element SHALL include the specified Code in the <code> element, except if the specified Code includes the keyword “SHOULD or “MAY”; in the latter case, this requirement is relaxed to the requirement strength of those keywords.

820 5. The header data element SHALL include the specified subsidiary Participation data elements in accordance with their specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1), using the specified Participation <typeCode> element, and in accordance with the specified Template ID and further constraints specified in the identified Technical Framework section.

825 The Header Content Module table may be supplemented with additional specific conformance requirements.

6.1.2.3.2 Entry Content Module Table

An Entry Content Module is specified using the following table.

Template ID					
Parent Template					
General Description					
Class/Mood	Code		Value Type	Value	
Opt	entryRelationship	Description	Template	Spec Document	Con-straint

This table implies the following conformance statements:

- 830
1. The entry SHALL include the specified Template ID in the <templateID> element of the clinical statement act element.
 2. The entry SHALL conform to all the requirements of the specified Parent Template.
 3. The entry SHALL include the specified classCode and moodCode values, and be conformant to the HL7 v3 requirements of that Act Class and Mood.

- 835 4. The entry SHALL include the specified entry Code in the <code> element of the clinical statement act element, except if the specified Section Code includes the keyword “SHOULD or “MAY”; in the latter case, this requirement is relaxed to the requirement strength of those keywords.
- 840 5. If of Class/Mood OBS/EVN, the entry SHALL include a value of the specified Data Type.
6. If Data Type is CD, the entry value SHALL be the specified Value.
- Note: The code may be specified as a value selected from an identified Value Set.
- 845 7. The entry SHALL include the specified subsidiary Entries in accordance with their specified Cardinality and Optionality (Opt column value, as described in Section 6.1.1), using the specified entryRelationship <typeCode> element, and in accordance with the specified Template ID and further constraints specified in the identified Technical Framework section.

The Entry Content Module table may be supplemented with additional specific conformance requirements.

850 **6.1.2.4 Value Sets**

Value sets, which are potentially reusable in a variety of contexts, are described separately from the content modules. Each value set is identified by name and OID, and its constituent concept values are listed in a table.

855 Value sets concepts may be drawn from multiple coding systems and some concepts may be represented in more than one coding system. When there is a choice of coding system, the content module that invokes the value set may establish constraints on when to use a particular system (e.g., based on local policy or national regulation). The content module that invokes the value set may also establish constraints on whether concepts not in the defined value set can be used (e.g., using the HL7 CWE [coded with exceptions] and CNE [coded no exceptions] domain 860 qualifiers); unless otherwise specified, the value set is extensible (CWE). The HL7 v3 CD data type allows the representation of a concept by a code together with a translation code in a different coding system; when multiple codes are provided for a concept, use of such translation codes is recommended.

865

6.2 CDA Release 2 Content Modules

6.2.1 Document Modules

6.2.1.1 Cardiac Imaging Report Specification 1.3.6.1.4.1.19376.1.4.1.1.1

870 This is the template for Cardiac Imaging Reports with discrete data elements as described in the ACC/AHA/et al. 2008 Key Data Elements and Definitions for Cardiac Imaging. The Template ID for conformance to this template is OID = 1.3.6.1.4.1.19376.1.4.1.1.1.

6.2.1.1.1 Format Code

The XSDDocumentEntry format code for this content is **urn:ihe:card:imaging:2011**

875 The mapping of CDA header attributes to XDS metadata shall be identical to the XDS-MS mapping specified in PCC TF-2: 4.1.1.

6.2.1.1.2 Parent Template

This document is a specialization of the IHE PCC Medical Document template (OID = 1.3.6.1.4.1.19376.1.5.3.1.1.1).

880 Note: The Medical Document includes requirements for various header elements; name, addr and telecom elements for identified persons and organizations; and basic participations record target, author, and legal authenticator.

This document is a specialization of the HL7 Procedure Note template (OID = 2.16.840.1.113883.10.20.18.1).

885 Note: This document is not a specialization of the HL7 Basic Diagnostic Imaging Report template due to conflicts with two Procedure Note requirements (format of serviceEvent/effectiveTime, and title on DICOM Catalogue section). When and if these are resolved, an instance may also comply to the Diagnostic Imaging Report.

6.2.1.1.3 Clinical Document Code

The ClinicalDocument/code for this document MAY be **18748-4 Diagnostic Imaging Study report 2.16.840.1.113883.6.1 LOINC STATIC**.

890 Note: Procedure Note specifies a list of alternate document codes. The selection of an appropriate document type code will depend on the intended access workflow, and may vary by local policy. As a query key, the code should allow efficient access to all relevant reports, while minimizing the number of query key values from which a user must choose. The LOINC code identified here was selected under the assumption that a clinical user looking for a cardiac imaging report would typically be interested in seeing all diagnostic imaging reports for the patient.

Additional codes may be found in CDA-PN.

895 6.2.1.1.4 Standards

Table 6.2.1.1.4

KDECI	ACC/AHA/et al. 2008 Key Data Elements and Definitions for Cardiac Imaging	http://content.onlinejacc.org/cgi/content/full/53/1/91
-------	---	---

CDAR2	HL7 CDA Release 2.0	http://www.hl7.org/documentcenter/private/standards/cda/r2/cda_r2_normativewebedition2010.zip
CDA-PN	HL7 Implementation Guide for CDA Release 2: Procedure Note (Universal Realm) (DSTU)	http://www.hl7.org/documentcenter/public/standards/dstu/CDAR2_IG_PROCNOTE_DSTU_R1_2010JUL.zip
CDA-DIR	Implementation Guide for CDA Release 2: Imaging Integration: Basic Imaging Reports in CDA and DICOM Diagnostic Imaging Reports (DIR) – Universal Realm (Informative)	http://www.hl7.org/documentcenter/private/standards/cda/igs/cdar2_ii_bimgrpts_r1_inf_2009mar.zip
CCD	ASTM/HL7 Continuity of Care Document Implementation Guide for CDAR2	http://www.hl7.org/documentcenter/private/standards/cda/igs/HL7_CCD_final.zip
PCC TF-2	IHE PCC Technical Framework Volume 2 Content Modules	http://www.ihe.net/Technical_Framework/upload/IHE_PCC_TF_Rev6-0_Vol_2_2010-08-30.pdf
DICOM	NEMA PS3.16 – DICOM Part 16: Content Mapping Resource	ftp://medical.nema.org/medical/dicom/2009/09_16pu.pdf

6.2.1.1.5 Data Element Requirements

900 Table 6.2.1.1.5-1 presents a summary of the data element requirements of KDECI, and their general mapping to sections in referenced CDA Implementation Guides. The data elements are summarized by category, and the number of discrete data elements in each category is listed. Implementation Guide sections listed in boldface are required in their respective standards.

Table 6.2.1.1.5-1 Summary of Data Element Requirements

KDECI	CDA-DIR	CDA-PN	CCD	PCC TF Medical Document
	DICOM Object Catalog (5)			
Administrative Facility (5) Data Source (1) Priority (1) Accreditation (2) Insurance (1)	CDA Header General (10) Document (19) Participants (20) Order (1) Service Event (12) Encounter (10)	CDA Header	CDA Header	CDA Header
Demographics (5)			Payers	
Study Referral Data (2)	Request			
History and Risk Factors	History	Medical History	Problems	

IHE Cardiology Technical Framework Supplement – Cardiac Imaging Report Content (CIRC)

KDECI	CDA-DIR	CDA-PN	CCD	PCC TF Medical Document
Vital Signs (4) Labs (2) Problems (14) Chest Pain (5) Family History (1) Tobacco Use (1) Risk Estimates (6) Functional assessment (2) Prior Tests (2) Prior Procedures (4) Medications (2) Contrast Reaction (1)		History of Present Illness		
		Social History	Social History	
		Family History	Family History	
		Procedure History	Procedures	
	Patient Presentation	Physical Examination		
		Vital Signs	Vital Signs	
			Functional status	
	Previous Findings		Results	
		Medication History	Medications	
		Allergies	Alerts	
	Study Description Identification (2) Indication (2) Physician (3) Procedure (1) Acquisition Parameters (14) Stress Method (4)		Planned Procedure	
Current Procedure Descriptions		Procedure Description		
Indications for Procedure		Procedure Indications		
		Medications Administered		
	Complications	Complications		
Study Findings—Ischemic Heart Disease (25)	Findings	Procedure Findings		
Study Findings—LV Function (6)				
Study Findings—Cardiac Morphology Chambers (6) Myocardium (5) Pulmonary Veins (1) Intracardiac Mass (1) Intracardiac Shunt (1) Pericardium (4) Valves (15) Aorta (3)				
Study Findings—Summary (6)	Summary	Postprocedure Diagnosis		
	Conclusions	Assessment and/or Plan		
	Recommendations			
	Key Images			

905 The data elements in KDECI are specified only as named concepts, with value sets as lists of terms. This document template (including its subsidiary section and entry templates) maps those concepts to appropriate coded concepts from SNOMED CT, LOINC, DICOM, or NDF-RT, using existing value sets from DICOM where applicable.

910 Note: In DICOM, value sets are known as Context Groups, and are identified by a Context Group ID (CID) as well by as an OID.

The “Recommended” data element requirement from KDECI is generally interpreted as a Required (R) data element optionality constraint in this profile (see Section 6.1.1.2).

6.2.1.1.6 Specification

915 This section specifies lower level content modules comprising the Cardiac Imaging Report, using Template ID as the key identifier.

Sections that are used according to the definitions in other specifications are identified with the relevant specification document. Additional constraints on vocabulary value sets, not specifically constrained within the section template, are also identified.

920 **Table 6.2.1.1.6-1 Cardiac Imaging Report Content Specification**

Template ID	1.3.6.1.4.1.19376.1.4.1.1.1			
Parent Template	Medical Document 1.3.6.1.4.1.19376.1.5.3.1.1.1 [PCC TF-2] Procedure Note 2.16.840.1.113883.10.20.18.1 [CDA-PN]			
General Description	<p>The Cardiac Imaging Report Content template specifies the content structure for a clinical report of a cardiology imaging exam, recorded in a DICOM Study. Such exams include:</p> <ul style="list-style-type: none"> • Echocardiography (transthoracic – TTE, transesophageal – TEE, and TTE stress) • Cardiac computed tomography (angiography – CTA, and coronary artery calcium scoring – CACS) • Cardiac magnetic resonance (angiography – MRA, and MR stress) • Cardiovascular nuclear medicine (SPECT myocardial perfusion, positron emission tomography – PET) • Diagnostic coronary catheter based fluoroscopy (interventional coronary angiography – ICA, and left ventriculography – LVG) <p>The template uses discrete data elements as described in the ACC/AHA/et al. 2008 Key Data Elements and Definitions for Cardiac Imaging.</p>			
Document Code	MAY be 18748-4, LOINC, “Diagnostic Imaging Study report”			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Header Elements				
M [1..1]	Encompassing Encounter	1.3.6.1.4.1.19376.1.4.1.3.1	CARD TF-2 6.2.3.1	
R [0..1]	Order	1.3.6.1.4.1.19376.1.4.1.3.2	CARD TF-2 6.2.3.2	

IHE Cardiology Technical Framework Supplement – Cardiac Imaging Report Content (CIRC)

M [1..1]	Patient Demographics	1.3.6.1.4.1.19376.1.4.1.3.3	CARD TF-2 6.2.3.3	
M [1..1]	Service Event and Performer	1.3.6.1.4.1.19376.1.4.1.3.4	CARD TF-2 6.2.3.4	CARD TF-2 6.2.1.1.6.1
M [1..1]	Legal Authenticator	1.3.6.1.4.1.19376.1.4.1.3.5	CARD TF-2 6.2.3.5	
Sections				
M [1..1]	Medical History (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.1	CARD TF-2 6.2.2.1	
M [1..1]	Medications	1.3.6.1.4.1.19376.1.5.3.1.3.19	PCC TF-2	CARD TF-2 6.2.1.1.6.2
M [1..1]	Allergies and Other Adverse Reactions	1.3.6.1.4.1.19376.1.5.3.1.3.13	PCC TF-2	CARD TF-2 6.2.1.1.6.3
R [1..1]	Coded Social History	1.3.6.1.4.1.19376.1.5.3.1.3.16.1	PCC TF-2	CARD TF-2 6.2.1.1.6.4
R [1..1]	Family History (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.5	CARD TF-2 6.2.2.2	
O [0..1]	Physical Examination	2.16.840.1.113883.10.20.2.10	CDA-PN	
R [1..1]	Vital Signs	1.3.6.1.4.1.19376.1.5.3.1.3.25	PCC TF-2	
R [1..1]	Prior Results (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.6	CARD TF-2 6.2.2.3	
O [0..1]	Cardiac Risk Assessment	1.3.6.1.4.1.19376.1.4.1.2.7	CARD TF-2 6.2.2.4	
M [1..1]	Indications and Planned Procedure	1.3.6.1.4.1.19376.1.4.1.2.8	CARD TF-2 6.2.2.5	CARD TF-2 6.2.1.1.6.5
R [1..1]	Cardiac Procedure Description	1.3.6.1.4.1.19376.1.4.1.2.9	CARD TF-2 6.2.2.6	
O [0..1]	Medications Administered	2.16.840.1.113883.10.20.18.2.8	CDA-PN	CARD TF-2 6.2.1.1.6.6
R [1..1]	Procedure Findings (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.11	CARD TF-2 6.2.2.7	
R [1..1]	Complications / Adverse Events	2.16.840.1.113883.10.20.18.2.4	CDA-PN	
R [1..1]	Postprocedure Diagnosis	2.16.840.1.113883.10.20.18.2.3	CDA-PN	
R [1..1]	Assessment And Plan (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.12	CARD TF-2 6.2.2.8	
O [0..1]	Key Images	1.3.6.1.4.1.19376.1.4.1.2.14	CARD TF-2 6.2.2.10	
C [1..1]	DICOM Object Catalog	1.3.6.1.4.1.19376.1.4.1.2.15	CARD TF-2 6.2.2.11	CARD TF-2 6.2.1.1.6.7
O [0..1]	Payer	1.3.6.1.4.1.19376.1.5.3.1.1.5.3.7	PCC TF-2	

6.2.1.1.6.1 Service Event Constraints

The value for serviceEvent / code SHOULD be drawn from value set [1.3.6.1.4.1.19376.1.4.1.5.2 Cardiac Imaging Procedures](#).

925 **6.2.1.1.6.2 Medications Section Constraints**

Within the Medications section the Content Creator SHALL be able to create a Medications entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.7 [PCC TF-2]) for each of the cardiac relevant medications identified in Value Set [1.3.6.1.4.1.19376.1.4.1.5.14 Cardiac Drug Classes](#), encoding the value in substanceAdministration/consumable/ManufacturedProduct/Material/code.

930 **6.2.1.1.6.3 Allergies and Other Adverse Reactions Section Constraints**

Within the Allergies and Other Adverse Reactions section the Content Creator SHALL be able to create an Allergies and Intolerances Concern Entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.5.3 [PCC TF-2]) for each of the cardiac imaging agent classes identified in Value Set [1.3.6.1.4.1.19376.1.4.1.5.10 Contrast Agents Classes for Adverse Reactions](#), encoding the value in observation/participant/participantRole/playingEntity/code.

6.2.1.1.6.4 Coded Social History Section Constraints

Within the Coded Social History section the Content Creator SHALL be able to create a Social History Observation Entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.13.4 [PCC TF-2]) with observation/code value {229819007, SNOMED CT, Tobacco use and exposure} and a CD data type observation/value from Value Set [1.2.840.10008.6.1.225 DICOM CID 3724 Smoking History](#).

Within the Coded Social History section the Content Creator SHALL be able to create a Social History Observation Entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.13.4 [PCC TF-2]) with observation/code value {363908000, SNOMED CT, Details of drug misuse behavior} and a CD data type observation/value of {78267003, SNOMED CT, Cocaine abuse}.

6.2.1.1.6.5 Indications and Planned Procedure Section Constraints

Within the Indications and Planned Procedure section the Content Creator SHALL be able to create a Problem Entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.5 [PCC TF-2]) for any of the cardiac relevant indications identified in Value Set [1.3.6.1.4.1.19376.1.4.1.5.27 Cardiac Imaging Procedure Indications](#).

6.2.1.1.6.6 Medications Administered Section Constraints

Within the Cardiac Procedure Description or the Medications Administered section the Content Creator SHALL be able to create a Medications entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.7 [PCC TF-2]) for each of the cardiac medication classes identified in Value Set [1.3.6.1.4.1.19376.1.4.1.5.15 Drug Classes Used in Cardiac Procedure](#), encoding the value in substanceAdministration/consumable/ManufacturedProduct/Material/code.

Within the Cardiac Procedure Description or the Medications Administered section the Content Creator SHALL be able to create a Medications entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.7 [PCC TF-2]) for the relevant cardiac contrast agents identified in Value Set

960 [1.3.6.1.4.1.19376.1.4.1.5.11 Contrast Agents](#), encoding the value in substanceAdministration/consumable/ManufacturedProduct/Material/code.

- Notes:
1. The set of contrast agents implemented may be limited to a subset of the Value Set, based on the types of procedures for which the Content Creator creates reports, hence the term “*relevant cardiac contrast agents*”.
 2. The Medications Administered section may be structured as a subsection of the Cardiac Procedure Description section.

965

6.2.1.1.6.7 DICOM Object Catalog Section Constraints

A DICOM Object Catalog Section SHALL be present if other document sections contain references to DICOM SOP Instances (images, structured report measurements, or other information objects), and MAY be present otherwise.

6.2.1.1.7 Conformance

970

CDA Release 2.0 documents that conform to the requirements of this document content module shall indicate their conformance by the inclusion of the <templateId> XML elements in the header of the document. A CDA Document may conform to more than one template. This content module inherits from the CDA-PN and the PCC TF Medical Document content modules, and so must conform to the requirements of those templates as well the Cardiac Imaging Report templates. This is shown in the sample document below, which includes <templateId> elements for the OIDs of all three templates.

975

```

980 <ClinicalDocument xmlns='urn:hl7-org:v3'>
  <typeId extension="POCD_HD000040" root="2.16.840.1.113883.1.3"/>
  <templateId root='1.3.6.1.4.1.19376.1.5.3.1.1.1' />
  <templateId root='1.3.6.1.4.1.19376.1.4.1.1.1' />
  <templateId root='2.16.840.1.113883.10.20.18.1' />
  <id root=' ' extension=' ' />
985 <code code='18748-4' displayName='Diagnostic Imaging Study report'
  codeSystem='2.16.840.1.113883.6.1' codeSystemName='LOINC' />
  <title>Cardiac Imaging Report</title>
  <effectiveTime value='20111004012005' />
  <confidentialityCode code='N' displayName='Normal'
990   codeSystem='2.16.840.1.113883.5.25' codeSystemName='Confidentiality' />
  <languageCode code='en-US' />
  :
  <component>
    <structuredBody>
995   :
    </structuredBody>
  </component>
</ClinicalDocument>

```

Figure 6.2.1.1.7-1 Cardiac Imaging Report Example Extract

1000

6.2.2 Section Modules

6.2.2.1 Medical History (Cardiac) Section

Template ID	1.3.6.1.4.1.19376.1.4.1.2.1			
Parent Template	Medical History 2.16.840.1.113883.10.20.18.2.5 [CDA-PN]			
General Description	The Medical History Section shall record past and current problems relevant to the patient’s cardiac presentation in the narrative block and in corresponding structured entries. Other non-cardiac medical history may also be recorded in the narrative, and may also be present in the entries.			
Section Code	11329-0, LOINC, “Medical History”			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Subsections				
O [0..1]	Active Problems	1.3.6.1.4.1.19376.1.5.3.1.3.6	PCC TF-2	
O [0..1]	History of Present Illness	1.3.6.1.4.1.19376.1.5.3.1.3.4	PCC TF-2	
O [0..1]	History of Past Illness	2.16.840.1.113883.10.20.2.9	CDA-PN	
Entries				
C [1..*]	Problem Concern Entry	1.3.6.1.4.1.19376.1.5.3.1.4.5.2	PCC TF-2	CARD TF-2 6.2.2.1.1
C [1..*]	Simple Observation	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.2.1.2

1005 The Medical History Section SHALL contain at least one Problem Concern Entry or at least one Simple Observation.

Note: Problems MAY be recorded directly in the Medical History Section, or in one or more subsections such as Active Problems, History of Present Illness, or History of Past Illness.

6.2.2.1.1 Problem Concern Entry Constraints

1010 A Content Creator SHALL be able to include a Problem Concern Entry for each of the conditions identified in Value Set [1.3.6.1.4.1.19376.1.4.1.5.4 Cardiac Problems/Concerns](#), encoding the value in act/entryRelationship/observation/code.

A Problem Concern Entry for {73211009, SNOMED CT, diabetes} SHALL use the specialized Diabetes Problem Entry (OID = 1.3.6.1.4.1.19376.1.4.1.4.1).

1015 A Problem Concern Entry for {194828000, SNOMED CT, angina} SHALL use the specialized Angina Problem Entry (OID = 1.3.6.1.4.1.19376.1.4.1.4.2).

6.2.2.1.2 Simple Observation Entry Constraints

A Content Creator SHALL be able to include a Simple Observation Entry for each of the conditions listed in value set [1.3.6.1.4.1.19376.1.4.1.5.4 Cardiac Problems/Concerns](#) indicating

1020 absence of the condition for the patient, represented with a compositional SNOMED CT code with qualifier [408729009 | finding context | = 410516002 | known absent].

Note: For example, absence of angina is thus represented by code 194828000:408729009=410516002. Such negative findings are represented with Simple Observations, and are not observations subsidiary to a Problem Concern act entry.

1025 A Content Creator SHALL be able to include a Simple Observation Entry with code {160245001, SNOMED CT, “No current problems or disability”}.

Note: Use of this “No current problems or disability” entry is mutually exclusive with Problem Concern entries.

6.2.2.2 Family History (Cardiac) Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.5		
Parent Template		Family History 1.3.6.1.4.1.19376.1.5.3.1.3.14 [PCC TF-2]		
General Description		The Family History (Cardiac) section shall include narrative and entries for family history of cardiac diseases.		
Section Code		10157-6, LOINC, “History of Family Member Diseases”		
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Entries				
M [1..*]	Simple Observation	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.2.1

6.2.2.2.1 Simple Observation Entry Constraints

1030 A Content Creator SHALL be able to create a Simple Observation Entry for each of the conditions listed in value set [1.2.840.10008.6.1.253 DICOM CID 3758 Cardiovascular Family History](#).

1035 Notes: 1. These observations are organized by disease, not by family member; hence, this section template is not a child of Coded Family Medical History Section 1.3.6.1.4.1.19376.1.5.3.1.3.15 [PCC TF-2], which requires a person-based history.

2. The specified Value Set includes concepts for “no family history of ..” and “unknown family history”.

6.2.2.3 Prior Results (Cardiac) Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.6		
Parent Template		Prior Results 1.3.6.1.4.1.19376.1.5.3.1.3.28 [PCC TF-2]		
General Description		The results section shall contain a narrative description of the relevant diagnostic procedures the patient received in the past, including procedures (studies) used for comparison. It shall include entries for procedures and references to procedure reports when known as described in the Entry Content Modules.		
Section Code		30954-2, LOINC, “Relevant Diagnostic Tests And/Or Laboratory Data”		

Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Entries				
O [0..*]	Procedure Entry	1.3.6.1.4.1.19376.1.5.3.1.4.19	PCC TF-2	CARD TF-2 6.2.2.3.1
O [0..*]	External References Entry	1.3.6.1.4.1.19376.1.5.3.1.4.4	PCC TF-2	
O [0..*]	Simple Observation	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.2.3.2

1040 The External References Entry allows reference to other documents, not necessarily linked to a specific identified procedure in a Procedure Entry.

6.2.2.3.1 Procedure Entry Constraints

1045 Procedure Entries SHOULD use codes for cardiac related imaging procedures from Value Set [1.3.6.1.4.1.19376.1.4.1.5.2 Cardiac Imaging Procedures](#) in the procedure/code element. The entry MAY include an actRelationship to an ExternalDocument act that is the report of the procedure.

The Procedure Entry of a study used for comparison (see Assessment and Plan (Cardiac) Section) SHALL include an ID attribute as a target for cross-reference.

6.2.2.3.2 Simple Observation Entry Constraints

1050 Simple Observation Entries MAY use codes for cardiac related lab results from Value Set [1.3.6.1.4.1.19376.1.4.1.5.13 Cardiac Lab Results](#) in the observation/code element.

Note: Depending on the method used for the lab test, a different LOINC code may be applicable, even though the test name is identical.

1055 6.2.2.4 Cardiac Risk Assessment Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.7		
Parent Template		none		
General Description		The Cardiac Risk Assessment Section shall record risk calculations made prior to the procedure in the narrative block and in corresponding entries.		
Section Code		X-CARDRISK, LOINC, "Assessment for risk of cardiovascular disease"		
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Entries				
R [1..*]	Simple Observation	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.2.4.1

6.2.2.4.1 Simple Observation Entry Constraints

The section SHALL contain a Simple Observation Entry for each cardiac risk estimate or score reported. Example risk assessment codes are listed in value set [1.3.6.1.4.1.19376.1.4.1.5.26 Cardiac Risk Scores](#).

1060

6.2.2.5 Indications and Planned Procedure Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.8		
Parent Template		Procedure Indications 2.16.840.1.113883.10.20.18.2.1 [CDA-PN]		
General Description		The Coded Procedure Indications section records details about the reason for the procedure. This section may include the pre-procedure diagnosis or diagnoses as well as one or more symptoms that contribute to the reason the procedure is being performed. Coded entries are required.		
Section Code		59768-2, LOINC, "Procedure Indications"		
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Subsections				
O [0..1]	Planned Procedure	2.16.840.1.113883.10.20.18.2.6	CDA-PN	
Entries				
R [1..*]	Problem Entry	1.3.6.1.4.1.19376.1.5.3.1.4.5	PCC TF-2	
O [0..*]	Planned Procedure Activity	1.3.6.1.4.1.19376.1.4.1.4.8	CARD TF-2: 6.2.4.7	

The planned investigations of the current procedure, based on the clinical indications, MAY be recorded in Planned Procedure Activity entries directly in the Indications and Planned Procedure Section, or in a subsidiary Planned Procedure subsection.

The Problem Entries recording the indications for the procedure MAY be recorded directly in the Indications and Planned Procedure Section, or as reason observations subsidiary to a Planned Procedure Activity entry.

Note: Clinicians often do not clearly distinguish between the indications for a procedure (the patient condition that justifies its performance) and the investigation to be done by the procedure that further evaluates that indication. The Indications and Planned Procedure Section allows both the problem and the planned procedure to be recorded together at the beginning of the exam, similar to the Assessment and Plan Section that allows the diagnosis from the procedure and the associated care plan to be recorded together at the end of the exam.

6.2.2.6 Cardiac Procedure Description Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.9		
Parent Template		Procedure Description 2.16.840.1.113883.10.20.18.2.2 [CDA-PN]		
General Description		The Cardiac Procedure Description section shall include narrative and entries identifying the procedure performed and the details of the procedure such as contrast agent(s) used,		

	imaging procedure protocol used, characteristics of the imaging equipment used, and medications administered during the exam.			
Section Code	29554-3, LOINC, "Procedure Description"			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Subsections				
O [0..1]	Medications administered	2.16.840.1.113883.10.20.18.2.8	CDA-PN	CARD TF-2: 6.2.2.6.2
Entries				
M [1..1]	Procedure Entry	1.3.6.1.4.1.19376.1.5.3.1.4.19	PCC TF-2	CARD TF-2: 6.2.2.6.1
C [1..*]	Medication (medications administered)	1.3.6.1.4.1.19376.1.5.3.1.4.7	PCC TF-2	CARD TF-2: 6.2.2.6.2
C [1..*]	Medication (contrast/imaging agents)	1.3.6.1.4.1.19376.1.5.3.1.4.7	PCC TF-2	CARD TF-2: 6.2.2.6.3
R [1..*]	Simple Observation (procedure protocol)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.6.4

Note: CDA conformance rules allow organization of content into implementation-defined subsections. For combined procedures (e.g., CACS and CCTA), each constituent procedure might be described in a subsection of Procedure Description. The conformance rules apply to any child subsections and entries of the main section.

1080 **6.2.2.6.1 Procedure Entry Constraints**

The Procedure Entry SHOULD use codes for cardiac related imaging procedures from Value Set [1.3.6.1.4.1.19376.1.4.1.5.2 Cardiac Imaging Procedures](#) in the procedure/code element.

6.2.2.6.2 Medication (medications administered) Constraints

1085 If medications other than contrast were administered, they SHALL be documented in a Medication Entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.7 [PCC TF-2]).

Note: The Medication Entry may appear either directly in the Procedure Description Section or in a subsidiary Medications Administered subsection or in a constituent procedure subsection.

1090 The Medication Entry SHOULD use the values of Value Set [1.3.6.1.4.1.19376.1.4.1.5.15 Drug Classes Used in Cardiac Procedure](#), with either the SNOMED CT or the NDF-RT coding system as required by local use, encoding the value in substanceAdministration/consumable/ManufacturedProduct/Material/code.

6.2.2.6.3 Medication (contrast/imaging agents) Constraints

If contrast/imaging agents were used, they SHALL be documented in a Medication Entry.

1095 Note: The Medication Activity Entry may appear either directly in the Procedure Description Section or in a subsidiary Medications Administered subsection or in a constituent procedure subsection.

The Medication Entry SHOULD use the values of Value Set [1.3.6.1.4.1.19376.1.4.1.5.11 Contrast Agents](#), encoding the value in substanceAdministration/consumable/ManufacturedProduct/Material/code.

Administered contrast agent volume or dose SHOULD be included in the entry in substanceAdministration /doseQuantity in accordance with the Medication Entry template.

6.2.2.6.4 Simple Observation (procedure protocol) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
C [1..1]	R: nuclear cardiology, CCTA, or CMR	RID11248, RadLex, “Cardiac Gating”	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.28 Cardiac Synchronization Techniques
C [1..1]	R: nuclear cardiology	122713, DCM, “Attenuation correction”	CD	n/a	1.2.840.10008.6.1.744 DICOM CID 3112 Attenuation Correction
C [1..1]	R: TTE, TEE	125203, DCM, “Acquisition Protocol”	CD	n/a	1.2.840.10008.6.1.616 DICOM CID 12224 Ultrasound Image Modes
C [1..1]	R: CMR	125203, DCM, “Acquisition Protocol”	ED (text/plain) or CD	n/a	Example values <ul style="list-style-type: none"> • Morphology and function • Delayed enhancement • Flow/velocity quantification • MR angiography • Perfusion • Other May use CD with value <ul style="list-style-type: none"> • {419997008, SNOMED CT, MR Angiography} • {419535008, SNOMED CT, MR Cardiac Perfusion}
C [1..1]	R: CCTA	8867-4, LOINC, Heart Rate	PQ	{HB}/min	n/a
C [1..1]	R: CCTA	RID12651, RadLex, “Number of CT detector rows”	PQ	{rows}	
C [1..1]	R: CCTA	RID12380, RadLex, “Gantry revolution time”	PQ	ms	
C [1..1]	R: CCTA	113813, DCM, “CT Dose Length Product Total”	PQ	mGy.cm	
C [1..1]	R: ICA, LVG	113722, DCM, “Dose Area Product Total”	PQ	Gy.m2	
C [1..1]	R: ICA, LVG	113730, DCM, “Total Fluoro Time”	PQ	s	
C [1..1]	R: stress CMR, stress TTE, stress SPECT, stress PET	121058, DCM, “Procedure reported”	CD		1.2.840.10008.6.1.755 DICOM CID 3200 Stress Test Procedure
C [1..1]	R: stress CMR, stress TTE, stress SPECT, stress PET	109056, DCM, “Stress Protocol”	CD		1.2.840.10008.6.1.57 DICOM CID 3261 Stress Protocols

C [1..1]	R: stress CMR, stress TTE, stress SPECT, stress PET	252130009, SNOMED CT, "Total Exercise duration"	PQ	min	
C [0..1]	R: stress CMR, stress TTE, stress SPECT, stress PET	246489000, SNOMED CT, "Pharmacological Stress Agent"	CD		1.2.840.10008.6.1.759 DICOM CID 3204 Stress Agents

6.2.2.7 Procedure Findings (Cardiac) Section

Template ID	1.3.6.1.4.1.19376.1.4.1.2.11				
Parent Template	Procedure Findings 2.16.840.1.113883.10.20.18.2.15 [CDA-PN]				
General Description	The Cardiac Procedure Findings section records clinically significant observations confirmed or discovered during the procedure.				
Section Code	59776-5, LOINC, "Procedure Findings"				
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint	
Subsections					
O [0..1]	Comparison to prior study (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.13	CARD TF-2: 6.2.2.9		
Entries					
R [1..*]	Simple Observation (ischemic heart disease)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.7.1	
C [1..1]	Coronary Angiography Arterial Analysis	1.3.6.1.4.1.19376.1.4.1.4.5	CARD TF-2: 6.2.4.5	CARD TF-2: 6.2.2.7.2	
R [1..*]	Simple Observation (LV function)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.7.3	
C [1..1]	Segmental Wall Analysis	1.3.6.1.4.1.19376.1.4.1.4.3	CARD TF-2: 6.2.4.4	CARD TF-2: 6.2.2.7.4	
R [1..*]	Simple Observation (cardiac morphology)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.7.5	

1105

Note: CDA conformance rules allow organization of content into implementation-defined subsections. For combined procedures (e.g., CACS and CCTA), each constituent procedure might be described in a subsection of Procedure Description. Alternatively, findings might be grouped into subsections by anatomy (e.g., a section for cardiac valve findings). The conformance rules apply to any child subsections and entries of the main section.

1110 The Procedure Findings (Cardiac) Section MAY include a subsection Comparison to Prior Study (Cardiac).

The Procedure Findings (Cardiac) Section and its sub-sections SHALL include Simple Observation Entries dependent on the type of procedure performed.

1115 Observation concepts for which the observation was not performed (or not assessed) MAY be represented in a Simple Observation Entry with a value using an HL7 Null Flavor. Explicit observations of conditions known to not be present SHALL be represented with a compositional SNOMED CT code with qualifier [408729009 | finding context | = 410516002 | known absent].

6.2.2.7.1 Simple Observation (ischemic heart disease) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
O [0..1]		102592004, SNOMED CT, "ECG Finding"	CD		164918000:246090004=128974000, SNOMED CT, "pathologic Q waves at baseline"
C [1..1]	R: CCTA, SPECT, CMR	102592004, SNOMED CT, "ECG Finding"	CD		1.3.6.1.4.1.19376.1.4.1.5.17 Rhythm Findings
O [0..1]		102592004, SNOMED CT, "ECG Finding"	CD		26141007:246090004=128974000, SNOMED CT, "ST-segment depression at baseline"
C [1..*]	R: stress TTE, stress SPECT, stress PET, CCTA, stress CMR	102592004, SNOMED CT, "ECG Finding"	CD		1.3.6.1.4.1.19376.1.4.1.5.18 ECG Findings
O [0..1]		122717, DCM, "Peak activity workload"	PQ	[MET]	
C [1..1]	R: stress TTE, stress SPECT	444981005, SNOMED CT, "Resting heart rate"	PQ	{HB}/min	
C [1..1]	R: stress TTE, stress SPECT	428420003, SNOMED CT, "Target heart rate"	PQ	{HB}/min	
C [1..1]	R: stress TTE, stress SPECT	428630002, SNOMED CT, "Maximum HR Achieved"	PQ	{HB}/min	
C [1..1]	R: stress TTE, stress SPECT	271649006:246090004=128974000, SNOMED CT, "Baseline systolic blood pressure"	PQ	mm[Hg]	
C [1..1]	R: stress TTE, stress SPECT	271650006:246090004=128974000, SNOMED CT, "Baseline diastolic blood pressure"	PQ	mm[Hg]	
C [1..1]	R: stress TTE, stress SPECT	314439003, SNOMED CT, "Maximum systolic blood pressure"	PQ	mm[Hg]	
C [1..1]	R: stress TTE, stress SPECT	314452008, SNOMED CT, "Maximum diastolic blood pressure"	PQ	mm[Hg]	
C [1..1]	R: exercise stress testing	122718, DCM, "Peak Double Product"	PQ	mm[Hg]. {HB}/min	
C [1..1]	R: exercise SPECT, exercise TTE, stress PET	418799008, SNOMED CT, "Symptom"	ED (text/plain)		Chest pain during exercise <ul style="list-style-type: none"> • Limiting chest pain • Nonlimiting chest pain • Anginal equivalent
O [0..1]		102592004, SNOMED CT, "ECG Finding"	CD		26141007, SNOMED CT, "ST-segment depression"

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
O [0..1]		429622005, SNOMED CT, "ST depression"	PQ	mv	
C [1..1]	R: CACS	112058, DCM, "Calcium Score" methodCode= 112055, DCM, Agatston Scoring Method	PQ	[arb'U]	

Adapted from KDECI Table 6

1120

6.2.2.7.2 Coronary Angiography Arterial Analysis Entry Constraints

The Coronary Angiography Arterial Analysis Entry SHALL be present if the procedure was a CCTA or ICA.

6.2.2.7.3 Simple Observation (LV function) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
R [1..1]		404684003, SNOMED CT, "Finding"	ED (text/plain) or CD		overall assessment of LV diastolic function. <ul style="list-style-type: none"> • Normal for age • Abnormal for age May use CD with value 395704004, SNOMED CT, "LV diastolic dysfunction" for finding of abnormal
R [1..1]		10230-1, LOINC, "LVEF" methodCode= <ul style="list-style-type: none"> • 258083009, SNOMED CT, "Visual estimation" • 258090004, SNOMED CT, "Calculated" 	PQ or IVL<PQ>	%	

1125

Adapted from KDECI Table 7

The LVEF numeric observation/value MAY be a categorical assessment using the HL7v3 Interval of Physical Quantities data type (IVL<PQ>), with preferred value ranges as shown in Table 6.2.2.7.3-1.

1130

Table 6.2.2.7.3-1 Preferred LVEF categorical ranges

Preferred LVEF categorical ranges
<30 %
30-40 %

Preferred LVEF categorical ranges
40-50 %
50-70 %
>70 %

6.2.2.7.4 Segmental Wall Analysis Entry Constraints

1135 The Segmental Wall Analysis Entry SHALL be present if the procedure was an ICA with left ventriculogram, stress SPECT, stress PET, stress TTE, or stress CMR.

For cath exams, the wall analysis observation targetSiteValue Set SHALL be [1.2.840.10008.6.1.219 DICOM CID 3718 Myocardial Wall Segments in Projection](#).

For stress exams, the wall analysis observation targetSiteValue Set SHALL be [1.2.840.10008.6.1.218 DICOM CID 3717 Myocardial Wall Segments](#).

1140

6.2.2.7.5 Simple Observation (cardiac morphology) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
R [1..1]		250929008, SNOMED-CT, left ventricular cavity size	CD		1.3.6.1.4.1.19376.1.4.1.5.22 Cardiac Chamber Size Assessments
O [0..1]		8823-7, LOINC, left ventricle systolic volume	PQ	ml	
O [0..1]		8821-1, LOINC, Left ventricle diastolic volume	PQ	ml	
C [0..1]	R: TTE, TEE, CMR O: CCTA	250964004, SNOMED-CT, right ventricular cavity size	CD		1.3.6.1.4.1.19376.1.4.1.5.22 Cardiac Chamber Size Assessments
C [0..1]	R: TTE, TEE, CMR O: CCTA	399121005, SNOMED-CT, Left atrium cavity size	CD		1.3.6.1.4.1.19376.1.4.1.5.22 Cardiac Chamber Size Assessments
C [0..1]	R: TTE, TEE, CMR O: CCTA	439749006:363698007=73829009, SNOMED-CT, Right atrium volume by imaging	CD		1.3.6.1.4.1.19376.1.4.1.5.22 Cardiac Chamber Size Assessments
O [0..1]	R: TTE, TEE	18154-5, LOINC, Interventricular septum Thickness diastole by US	PQ	mm	
O [0..1]	R: TTE, TEE	18152-9, LOINC, Left Ventricle Posterior Wall Diastolic Thickness	PQ	mm	
O [0..1]	R: TTE, TEE	18153-7, LOINC, Right Ventricular Anterior Wall Diastolic Thickness	PQ	mm	

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
O [0..1]		18087-7, LOINC, Left Ventricle Mass	CD		260395002, SNOMED CT, “normal” 35105006, SNOMED CT, “Increased”
O [0..1]	R: CMR, CCTA	304522008, SNOMED CT, Pulmonary vein finding	CD		1.3.6.1.4.1.19376.1.4.1.5.23 Pulmonary Veins Assessments
O [0..1]		404684003, SNOMED CT, “Finding”	ED (text/plain) or CD		indicate the type of intracardiac mass if present. <ul style="list-style-type: none"> • Vegetation • Thrombus • Neoplasm • Mass of Unknown Etiology May use CD with value <ul style="list-style-type: none"> • 387842002, SNOMED CT, “neoplasm of heart” • 309519009, SNOMED CT, “LV Thrombus”
O [0..*]		442119001, SNOMED CT, “Cardiac shunt finding”			1.3.6.1.4.1.19376.1.4.1.5.29 Cardiac Shunt Types
C [1..1]	R: TTE, TEE, CMR, CCTA, ICA, LVG	301123005, SNOMED CT, “Pericardial finding”	CD		373945007, SNOMED CT, “Pericardial effusion” + size [CARD TF-2: 6.2.2.7.5.1]
O [0..1]		301123005, SNOMED CT, “Pericardial finding”	CD		35304003, SNOMED CT, “Tamponade”
O [0..1]		301123005, SNOMED CT, “Pericardial finding”	ED text/plain or CD		Indicate the thickness of the pericardium. <ul style="list-style-type: none"> • Normal • Thickened • Calcified May use CD with value 42653000, SNOMED CT, “Calcified pericardium”
C [1..1]	R: TTE, TEE, CCTA, CMR	301099004, SNOMED CT, “Aortic valve finding”	CD		301100007, SNOMED CT, “Aortic valve normal” 84683006, SNOMED CT, “Aortic valve prosthesis” 8722008, SNOMED CT, “Aortic valve disorder”
O [0..*]		301099004, SNOMED CT, “Aortic valve finding”	CD		253612007, SNOMED CT, aortic valve cusp prolapse 301184001, SNOMED CT, aortic valve vegetations 13689005, SNOMED CT, congenital anomaly of aortic

IHE Cardiology Technical Framework Supplement – Cardiac Imaging Report Content (CIRC)

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
					valve
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	301099004, SNOMED CT, “Aortic valve finding”	CD		60573004, SNOMED CT, aortic valve stenosis + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	301099004, SNOMED CT, “Aortic valve finding”	CD		60234000, SNOMED CT, Aortic regurgitation + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CCTA, CMR	301101006, SNOMED CT, “Mitral valve finding”	CD		301103009, SNOMED CT, “Mitral valve normal” 11851006, SNOMED CT, “Mitral valve disorder” 17107009, SNOMED CT, “Mitral valve prosthesis” 360063009, SNOMED CT, “Annuloplasty ring”
O [0..*]	R: TTE, TEE, CCTA, CMR	301101006, SNOMED CT, “Mitral valve finding”	CD		409712001, SNOMED CT, Mitral valve prolapse 270906004, SNOMED CT, mitral chordae rupture 301185000, SNOMED CT, Mitral valve vegetations 75372006, SNOMED CT, congenital anomaly of Mitral valve
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	301101006, SNOMED CT, “Mitral valve finding”	CD		251002009, SNOMED CT, mitral valve annular calcification
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	301101006, SNOMED CT, “Mitral valve finding”	CD		79619009, SNOMED CT, Mitral valve stenosis + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	301101006, SNOMED CT, “Mitral valve finding”	CD		48724000, SNOMED CT, Mitral regurgitation + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CMR	301108000, SNOMED CT, “Tricuspid valve finding”	CD		301111004, SNOMED CT, “Tricuspid valve normal” 20721001, SNOMED CT, “Tricuspid valve disorder” 25510005, SNOMED CT, “Cardiac valve prosthesis” 360063009, SNOMED CT,

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
					“Annuloplasty ring”
C [1..1]	R: TTE, TEE, CMR	301108000, SNOMED CT, “Tricuspid valve finding”	CD		49915006, SNOMED CT, Tricuspid valve stenosis + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CMR	301108000, SNOMED CT, “Tricuspid valve finding”	CD		111287006, SNOMED CT, Tricuspid regurgitation + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CMR	301104003, SNOMED CT, “Pulmonic valve finding”	CD		301107005, SNOMED CT, “Pulmonic valve normal” 76267008, SNOMED CT, “Pulmonic valve disorder” 423449009, SNOMED CT, “Pulmonary valve prosthesis”
C [1..1]	R: TTE, TEE, CMR	301104003, SNOMED CT, “Pulmonic valve finding”	CD		56786000, SNOMED CT, Pulmonic valve stenosis + severity [CARD TF-2: 6.2.2.7.5.2]
O [0..1]		301104003, SNOMED CT, “Pulmonic valve finding”	CD		91434003, SNOMED CT, Pulmonic regurgitation + severity [CARD TF-2: 6.2.2.7.5.2]
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	404684003, SNOMED CT, “Finding” + targetSiteCode [CARD TF-2: 6.2.2.7.5.3]	CD		308546005, SNOMED CT, “Dissection of aorta”
C [1..1]	R: TTE, TEE, CCTA, CMR, ICA, LVG	404684003, SNOMED CT, “Finding”	CD		251036003, SNOMED CT, “Aortic root dilation”

Adapted from KDECI Table 8

6.2.2.7.5.1 Pericardial Effusion Size Observation

1145 An observation of pericardial effusion MAY include a post-coordinated interpretation code from the values of Table 6.2.2.7.5-1.

Table 6.2.2.7.5-1 Pericardial Effusion post-coordinated codes

SNOMED CT expression	Meaning
373945007	Pericardial effusion
373945007:408729009=410516002	Pericardial effusion absent
373945007:363713009=371929004	Pericardial effusion – minimally significant
373945007:363713009=255507004	Pericardial effusion – small

SNOMED CT expression	Meaning
373945007:363713009=371927002	Pericardial effusion – moderate
373945007:363713009=255509001	Pericardial effusion – large

6.2.2.7.5.2 Severity Observation

1150 An observation of valve stenosis or valve regurgitation MAY include an optional entryRelationship element indicating the severity of the finding. When present, this entryRelationship element SHALL contain an observation conforming to the Severity entry (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.1 [PCC TF]). The entryRelationship typeCode SHALL be “SUBJ” and inversionInd SHALL be “true”.

6.2.2.7.5.3 Dissection targetSite Observation

1155 An observation of Dissection of aorta SHALL include a targetSiteCode of {54247002, SNOMED CT, “Ascending aorta”} for Stanford Type A dissections (involving the ascending aorta regardless of site of origin).

6.2.2.8 Assessment And Plan (Cardiac) Section

1160

Template ID	1.3.6.1.4.1.19376.1.4.1.2.12			
Parent Template	Assessment And Plan 2.16.840.1.113883.10.20.18.2.14 [CDA-PN]			
General Description	The assessment and plan section shall contain a narrative description of the assessment of the patient condition and expectations for care including proposals, goals, and order requests for monitoring, tracking, or improving the condition of the patient. This cardiac specialization requires coded entries for specific findings; and specifies use of the combined assessment and plan section, not the separate sections for assessment and for plan.			
Section Code	51847-2, LOINC, “Assessment and Plan”			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint
Subsections				
O [0..1]	Comparison to prior study (Cardiac)	1.3.6.1.4.1.19376.1.4.1.2.13	CARD TF-2: 6.2.2.9	
Entries				
R [1..*]	Simple Observation (summary)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.8.1
O [0..*]	Plan of Care Activities	2.16.840.1.113883.10.20.1.25	CCD	

The Assessment and Plan (Cardiac) Section MAY include a subsection Comparison to Prior Study (Cardiac).

6.2.2.8.1 Simple Observation (summary) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
C [1..1]	R: CCTA, ICA, LVG, stress SPECT, stress TTE	271925006:363702006=414 545008, SNOMED CT, “Ischemic heart disease assessment”	CD		1.3.6.1.4.1.19376.1.4.1.5.24 Imaging Findings
R [1..1]		250907009, SNOMED CT, “left ventricular function”	CD		1.3.6.1.4.1.19376.1.4.1.5.24 Imaging Findings
R [1..1]		271921002 SNOMED CT “ECG Finding”	CD		1.3.6.1.4.1.19376.1.4.1.5.30 ECG Summary Findings

1165

6.2.2.9 Comparison to Prior Study (Cardiac) Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.13			
Parent Template		none			
General Description		The Comparison to Prior Study (Cardiac) Section identifies a prior exam used for comparison, and describes any significant changes from that exam to the current exam.			
Section Code		111424, DICOM, “Comparison to Previous Exams”			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint	
Entries					
R [1..*]	Simple Observation (comparison)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.2.9.1	

1170

The Comparison to Prior Study (Cardiac) Section SHALL reference a prior study identified in the Prior Results (Cardiac) Section using the ID attribute in a <reference> element in section / text. The narrative SHALL include the date and type of the comparison exam.

6.2.2.9.1 Simple Observation (comparison) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
R [1..1]		111424, DCM, “Comparison to previous exams”	CD	n/a	1.2.840.10008.6.1.770 DICOM CID 3217 Comparison Finding
O [0..1]	R: Stress	248243004, SNOMED CT, “Exercise tolerance”	CD	n/a	1.2.840.10008.6.1.782 DICOM CID 3236 Tolerance Comparison Findings
O [0..1]		251053005, SNOMED CT, “Myocardial Perfusion”	CD	n/a	1.2.840.10008.6.1.781 DICOM CID 3235 Perfusion Comparison Findings

O [0..1]	R: TTE, TEE. LVG	250909007, SNOMED CT, “LV Wall motion”	CD	n/a	1.2.840.10008.6.1.783 DICOM CID 3237 Wall Motion Comparison Findings
----------	---------------------	---	----	-----	--

6.2.2.10 Key Images Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.14			
Parent Template		CDA-DIR: 3.1.1 (no Template OID)			
General Description		The Key Images section contains narrative description of and references to DICOM Image Information Objects that illustrate the findings of the procedure reported.			
Section Code		55113-5, LOINC, “Key Images”			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint	
Entries					
O [0..*]	SOP Instance Observation	2.16.840.1.113883.10.20.6.2.8	CDA-DIR		

1175 The Key Images section text SHALL contain image references using `linkHtml` elements, where `@href` is a valid Web Access to DICOM Persistent Object (WADO) URL and the text content of `linkHtml` is the either visible text of the hyperlink or a descriptor or identifier of the image.

1180 The Key Images section MAY include SOP Instance Observation entries equivalent to the `linkHtml` image references.

6.2.2.11 DICOM Object Catalog Section

Template ID		1.3.6.1.4.1.19376.1.4.1.2.15			
Parent Template		none			
General Description		DICOM Object Catalog lists all referenced objects and their parent Series and Studies, plus other DICOM attributes required for retrieving the objects.			
Section Code		121181, DCM, “DICOM Object Catalog”			
Opt	Data Element or Section Name	Template ID	Specification Document	Constraint	
Entries					
M [1..*]	Study Act	2.16.840.1.113883.10.20.6.2.6	CDA-DIR		

1185 The DICOM Object Catalog Section SHALL provide Study Act Entries, including subsidiary Series and SOPInstance Acts, that identify all SOP Instances referenced in other document

sections. It MAY provide Study Act Entries for the DICOM studies associated with the documented procedure, even if no SOP Instances are explicitly referenced.

The DICOM Object Catalog Section MAY include section text that indicates in human readable form that the content of the section is computer-processable links to DICOM information objects, and there is no other attestable report content.

1190

Note: CIRC and its parent Procedure Note require a section/title attribute for all document sections. CDA-DIR r1 forbids section/title for the DICOM Object Catalog Section, hence this template is distinct from the CDA-DIR DICOM Object Catalog Section template.

1195

6.2.3 Header Content Modules

6.2.3.1 Encompassing Encounter

Template ID		1.3.6.1.4.1.19376.1.4.1.3.1			
Parent Template		none			
General Description		The setting of the clinical encounter during which the documented act(s) or ServiceEvent occurred. This template includes specializations of Responsible Party, Health Care Facility, Referring Provider, and Physician of Record.			
Header Element		componentOf / encompassingEncounter			
Code		Not constrained			
Opt	Participation	Description	Template	Spec Document	Con-straint
O [0..1]	RESP	Responsible Party		CARD TF-2: 6.2.3.1.1	
R [1..1]	LOC	Health Care Facility		CARD TF-2: 6.2.3.1.2	
O [0..1]	REF	Referring Provider		CARD TF-2: 6.2.3.1.3	
C [0..1]	ATND	Physician of Record	2.16.840.1.113883.10.20.6.2.2	CDA-DIR	CARD TF-2: 6.2.3.1.4

The `encompassingEncounter / effectiveTime` value SHALL be accurate to the day, and MAY be accurate to the second.

1200

```

1205 <componentOf>
      <encompassingEncounter xmlns:ihecard='urn:ihe:card'>
        <templateId root='1.3.6.1.4.1.19376.1.4.1.3.1' />
        <effectiveTime value="20110407"/>
        <responsibleParty>
          <assignedEntity>
            <id root="2.16.840.1.113883.4.6" extension="physician NPI" />
            <code code=" " codeSystem="2.16.840.1.113883.6.101"
1210       codeSystemName="nuccProviderCodes" displayName=" " />
            <ihecard:accreditation accreditation descriptive string />
            <assignedPerson>
              <name>responsible physician name</name>
            </assignedPerson>
          </assignedEntity>
        </responsibleParty>
1215 <encounterParticipant typeCode="REF">
          <assignedEntity>
            <id root="2.16.840.1.113883.4.6" extension=" " />
            <code code=" " codeSystem="2.16.840.1.113883.6.101"
1220       codeSystemName="nuccProviderCodes" displayName=" " />
            <addr>referring physician address</addr>
            <telecom>referring physician phone</telecom >
            <assignedPerson>
              <name>referring physician name</name>
1225       </assignedPerson>
            </assignedEntity>
          </encounterParticipant>
          <location>
            <healthCareFacility>
1230       <id root=" " extension=" " />
              <code code="CARD" codeSystem="2.16.840.1.113883.5.111"
                codeSystemName="roleCode" displayName="Cardiology Clinic">
                <ihecard:accreditation accreditation descriptive string />
            <serviceProviderOrganization>
1235       <name>organization name</name>
              <addr>organization address</addr>
              <telecom>organization phone, web, other telecom access</telecom >
            </serviceProviderOrganization>
            <location>
1240       <name>service location name</name>
              <addr>service location address</addr>
            </location>
          </healthCareFacility>
        </location>
1245       ...
      </encompassingEncounter >
    </componentOf>

```

Figure 6.2.3.1-1 Encompassing Encounter Example

6.2.3.1.1 Responsible Party

1250 The responsible party element represents only the party responsible for the encounter, not necessarily the entire episode of care.

The responsibleParty element MAY be present. If present, responsibleParty/assignedEntity SHALL have at least one assignedPerson or representedOrganization element present.

1255 Note: This is identical to CDA-DIR CONF-DIR-67

responsibleParty assignedEntity id SHALL be present with the responsible physician's identifier.

`assignedEntity` code SHOULD be present with the responsible physician's specialty.

1260 `assignedEntity` MAY include an `accreditation` element from the **urn:ihe:card** namespace to provide physician accreditation status.

The `accreditation` element SHALL use the character string (ST) data type.

The `accreditation` element SHALL appear after the defined elements of the Role class, and before any `scoper` or `player` entity elements.

1265 `assignedEntity assignedPerson name` SHALL be present with the responsible physician's name.

6.2.3.1.2 Health Care Facility

Conformance requirements for encoding of information about the Health Care Facility are provided in the Procedure Note Implementation Guide, and are reproduced here for reference:

1270 **CONF-PN-18:** A Procedure Note SHOULD contain information about where the procedure was performed.

CONF-PN-19: If present, the physical location of the procedure SHALL be represented with `componentOf/encompassingEncounter/location/healthCareFacility/id` element.

1275 **CONF-PN-20:** If present, the `location/healthCareFacility` element SHALL contain a `code` element representing the type of location.

Copied from CDA-PN

`Location/healthCareFacility/serviceProviderOrganization` SHOULD be present with `name`, `addr` and `telecom` elements.

1280 `Location/healthCareFacility/location` (Place Entity) MAY be present with `name` and/or `addr` elements identifying the place of the encounter.

`healthCareFacility` MAY include an `accreditation` element from the **urn:ihe:card** namespace to provide facility accreditation status.

The `accreditation` element SHALL use the character string (ST) data type.

1285 The `accreditation` element SHALL appear after the defined elements of the Role class, and before any `scoper` or `player` entity elements.

6.2.3.1.3 Referring Provider

Conformance requirements for encoding of information about the referring provider are provided in the Procedure Note Implementation Guide, and are reproduced here for reference:

1290 **CONF-PN-21:** A Procedure Note MAY contain information about the referring provider.

CONF-PN-22: If present, the referring provider SHALL be represented with a `componentOf/encompassingEncounter/encounterParticipant` element.

1295 CONF-PN-23: When an `encompassingEncounter/encounterParticipant` representing the referring provider is present, the `encounterParticipant/@typeCode` SHALL be REF (referrer) and an `assignedEntity` SHALL be present.

CONF-PN-24: If elements required in `componentOf/encompassingEncounter/encounterParticipant` are unknown, these elements SHALL be represented with the appropriate HL7 null value.

Copied from CDA-PN

1300

If present, the referring provider `assignedEntity id` SHALL be present with one value of the set being the Performer's National Provider Identifier (root OID = 2.16.840.1.113883.4.6) (in the US) or the equivalent.

1305 The referring provider `assignedEntity code` SHOULD be present with the referring provider's specialty.

6.2.3.1.4 Physician of Record

The Physician of Record participation SHOULD be present if the procedure is performed in an in-patient context; it MAY be present otherwise.

1310 Conformance requirements for encoding of information about the Physician of Record are taken from the Diagnostic Imaging Report Implementation Guide.

CONF-DIR-68: A Physician of Record Participant (`templateId` 2.16.840.1.113883.10.20.6.2.2) SHOULD be present.

CONF-DIR-69: The `templateId` for a Physician of Record Participant SHALL be 2.16.840.1.113883.10.20.6.2.2.

1315 CONF-DIR-70: A Physician of Record Participant SHALL be represented with an `encounterParticipant` element where `@typeCode` is ATND.

CONF-DIR-71: An `encounterParticipant/assignedEntity/id` element SHALL be present containing the ID of the physician of record.

1320 CONF-DIR-72: An `encounterParticipant/assignedEntity/code` element SHALL be present and SHALL contain ... an appropriate national health care provider coding system (e.g., NUCC in the U.S., where `@codeSystem` is 2.16.840.1.113883.11.19465).

CONF-DIR-73: An `assignedPerson/name` element SHOULD be present containing the name of the physician of record.

Copied from CDA-DIR

1325

6.2.3.2 Order

Template ID	1.3.6.1.4.1.19376.1.4.1.3.2
Parent Template	none
General Description	Identification of orders that are fulfilled by this document.
Header Element	inFulfillmentOf / order
Code	Not constrained

The element `order / priorityCode` SHALL be present, with values from coding system HL7 ActPriority (OID = 2.16.840.1.113883.5.7).

1330 One value within the set of `order / id` values SHALL be the Accession Number used in the DICOM imaging data, with the root representing the Assigning Authority (Issuer of Accession Number).

```

1335 < inFulfillmentOf>
      <order>
1340   <templateId root='1.3.6.1.4.1.19376.1.4.1.3.2' />
      <id root="assigning authority" extension="accession number" />
      <code code=" " codeSystem=" "
        codeSystemName=" " displayName=" ">
      <priorityCode code=" " codeSystem=" 2.16.840.1.113883.5.7"
        codeSystemName="ActPriority" displayName=" ">
      </order>
</ inFulfillmentOf>
    
```

Figure 6.2.3.2-1 Order

1345 **6.2.3.3 Patient Demographics**

Template ID	1.3.6.1.4.1.19376.1.4.1.3.3
Parent Template	none
General Description	Identification of the patient demographics.
Header Element	recordTarget / patientRole
Code	n/a

The CDA Header SHALL include a Patient Entity.

1350 The CDA Header Patient Entity SHOULD include the elements `birthTime`, `administrativeGenderCode`, `raceCode` and `ethnicGroupCode`. Note that these elements may be encoded with a “flavor of null” if the information is unknown or collection is prohibited by local regulation.

```

1355 <recordTarget>
      <patientRole>
        <id root="2.16.840.1.113883.19.5" extension="12345"/>
        <patient>
          <name>Charles David Anderson II</name>
          <administrativeGenderCode code="M" codeSystem="2.16.840.1.113883.5.1"/>
          <birthTime value="19320924"/>
1360 <raceCode code=" 2076-8" codeSystem="2.16.840.1.113883.5.104"
          codeSystemName="Race" displayName="Native Hawaiian or Other Pacific Islander">
          <ethnicGroupCode code=" 2186-5" codeSystem="2.16.840.1.113883.5.50"
          codeSystemName="Ethnicity" displayName="Not Hispanic">
        </patient>
1365 <providerOrganization>
          <name>Maui Healthcare</name>
          <addr>Pukalani, Maui, HI<addr>
        </providerOrganization>
      </patientRole>
    </recordTarget>
  
```

Figure 6.2.3.3-1 Demographics

6.2.3.4 Service Event and Performer

Template ID		1.3.6.1.4.1.19376.1.4.1.3.4			
Parent Template		none			
General Description		Description of the procedure being documented			
Header Element		documentationOf/ serviceEvent			
Code		SHOULD be selected from code system 2.16.840.1.113883.6.96 SNOMED CT and MAY be selected from a localized procedure coding system for a given country such as 2.16.840.1.113883.6.104 ICD9 CM Procedures or 2.16.840.1.113883.6.12 CPT-4 in the U.S.			
Opt	Participation	Description	Template	Spec Document	Constraint
M [1..*]	PPRF	Primary Performer		CARD TF-2: 6.2.3.4	

1375 Conformance requirements for encoding of information about the Service Event and Performer are provided in the Procedure Note Implementation Guide, and are reproduced here for reference:

1380 CONF-PN-31: A Procedure Note SHALL contain one or more documentationOf/ serviceEvent elements.

CONF-PN-32: The value for serviceEvent/code SHOULD be selected from code system 2.16.840.1.113883.6.96 SNOMED CT and MAY be selected from a localized procedure coding system for a given country such as 2.16.840.1.113883.6.104 ICD9 CM Procedures or 2.16.840.1.113883.6.12 CPT-4 in the U.S.

CONF-PN-33: The serviceEvent/effectiveTime SHALL be present with effectiveTime/low and SHALL include effectiveTime/high if

1385 effectiveTime/width is not present. The serviceEvent/effectiveTime SHALL be accurate to the day, and MAY be accurate to the second.

CONF-PN-34: If the date and only the general length of the procedure are known, the serviceEvent/effectiveTime/low SHALL be present with an effectiveTime/width element. The serviceEvent/effectiveTime/low SHALL be accurate to the day, and MAY be accurate to the second.

1390 CONF-PN-35: If only the date is known and the duration of the procedure is unknown, the serviceEvent/effectiveTime/width element SHALL contain the appropriate HL7 null value.

CONF-PN-36: The primary performers (PPRF) SHALL be identified.

CONF-PN-37: For all performers, serviceEvent/performer/assignedEntity/code SHALL be present.

1395 CONF-PN-38: The value for serviceEvent/performer/assignedEntity/code SHOULD be selected from a localized assignedEntity coding system for a given country and MAY be selected from code system 2.16.840.1.113883.11.19465 Healthcare Provider Taxonomy Code (NUCC).

1400 CONF-PN-39: Any assistants SHALL be identified and SHALL be identified as secondary performers (SPRF).

Copied from CDA-PN

The value for serviceEvent code SHOULD be drawn from value set
1405 [1.3.6.1.4.1.19376.1.4.1.5.1 Cardiac Imaging Services](#).

The set of serviceEvent id values SHALL include the Study Instance UID used in the DICOM imaging data, with the UID value in the root attribute.

If the performer is a physician, assignedEntity id SHALL be present with the Performer's National Provider Identifier (root OID = 2.16.840.1.113883.4.6) (in the US) or the equivalent.

1410 assignedEntity MAY include an accreditation element from the **urn:ihe:card** namespace to provide physician accreditation status.

The accreditation element SHALL use the character string (ST) data type.

The accreditation element SHALL appear after the defined elements of the Role class, and before any scoper or player entity elements.

1415


```

1420 <performer typeCode=" ">
      <templateId root=' '/>
      <assignedEntity>
        <id root="2.16.840.1.113883.4.6" extension=" " />
        <code code=" " codeSystem="2.16.840.1.113883.6.101"
          codeSystemName="nuccProviderCodes" displayName=" ">
          <assignedPerson>
            <name>performer name</name>
          </assignedPerson>
1425 </assignedEntity>
      </performer>
    
```

Figure 6.2.3.4-1 Performer

6.2.3.5 Legal Authenticator

Template ID	1.3.6.1.4.1.19376.1.4.1.3.5
Parent Template	none
General Description	Description of the document’s legal authenticator, identifying the person signing the report
Header Element	legalAuthenticator / assignedEntity
Code	SHOULD be present with the Legal Authenticator’s specialty, with values from coding system NUCC Health Care Provider Taxonomy (in the US) or the equivalent

1430 One value within the set of `assignedEntity id` values SHOULD be the Legal Authenticator’s National Provider Identifier (in the US) or the equivalent.

6.2.4 Entry Content Modules

1435

6.2.4.1 Diabetes Problem Entry

Template ID		1.3.6.1.4.1.19376.1.4.1.4.1			
Parent Template		Problem Entry 1.3.6.1.4.1.19376.1.5.3.1.4.5 [PCC TF-2]			
General Description		The Diabetes Problem Entry is a problem observation with a required subsidiary entry that further describes the diabetic therapy.			
Class/Mood	Code	Value Type	Value		
OBS / EVN	404684003, SNOMED CT, “Finding”	CD	73211009, SNOMED CT, “Diabetes”		
Opt	entryRelationship	Description	Template	Spec Document	Con-straint
R [1..*]	RSON inversionInd = “true”	Diabetic Therapy	1.3.6.1.4.1.19376.1.4.1.4.1.1	CARD TF-2: 6.2.4.2	
O [0..1]	SUBJ inversionInd = “true”	Severity	1.3.6.1.4.1.19376.1.5.3.1.4.1	PCC TF-2	
O [0..1]	REFR	Problem Status Observation	1.3.6.1.4.1.19376.1.5.3.1.4.1.1	PCC TF-2	
O [0..1]	REFR	Health Status Observation	1.3.6.1.4.1.19376.1.5.3.1.4.1.2	PCC TF-2	
O [0..*]	SUBJ inversionInd = “true”	Comment	1.3.6.1.4.1.19376.1.5.3.1.4.2	PCC TF-2	

The Problem Entry template provides additional guidance regarding observation elements `templateId`, `id`, `text`, `statusCode`, `effectiveTime`, and `originalText`.

1440

6.2.4.2 Diabetes Therapy Entry

Template ID		1.3.6.1.4.1.19376.1.4.1.4.1.1			
Parent Template		none			
General Description		The Diabetes Therapy Entry is a procedure act that describes the therapy.			
Class/Mood	Code	Value Type	Value		
PROC / EVN	from 1.2.840.10008.6.1.223 DICOM CID 3722 Diabetic Therapy	n/a			

6.2.4.3 Angina Problem Entry

Template ID	1.3.6.1.4.1.19376.1.4.1.4.2				
Parent Template	Problem Entry 1.3.6.1.4.1.19376.1.5.3.1.4.5 [PCC TF-2]				
General Description	The Angina Problem Entry is a problem observation with subsidiary entries that further describe the problem.				
Class/Mood	Code		Value Type	Value	
OBS / EVN	404684003, SNOMED CT, “Finding”		CD	194828000, SNOMED CT, “Angina”	
Opt	entryRelationship	Description	Template ID	Spec Document	Con-straint
R [1..1]	COMP	Simple Observation (angina stability)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.4.3.1
O [0..*]	REFR	Simple Observation (angina characteristics)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2: 6.2.4.3.2
O [0..1]	SUBJ inversionInd = “true”	Severity	1.3.6.1.4.1.19376.1.5.3.1.4.1	PCC TF-2	
O [0..1]	REFR	Problem Status Observation	1.3.6.1.4.1.19376.1.5.3.1.4.1.1	PCC TF-2	
O [0..1]	REFR	Health Status Observation	1.3.6.1.4.1.19376.1.5.3.1.4.1.2	PCC TF-2	
O [0..*]	SUBJ inversionInd = “true”	Comment	1.3.6.1.4.1.19376.1.5.3.1.4.2	PCC TF-2	

1445 The Problem Entry template provides additional guidance regarding observation elements templateId, id, text, statusCode, effectiveTime, and originalText.

6.2.4.3.1 Simple Observation (angina stability) Constraint

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
R [1..1]		404684003, SNOMED CT, “Finding”	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.7 Angina Type

6.2.4.3.2 Simple Observation (angina characteristics) Constraint

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
O [1..*]		404684003, SNOMED CT, “Finding”	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.25 Angina Characteristics

1450

6.2.4.4 Segmental Wall Analysis

Template ID		1.3.6.1.4.1.19376.1.4.1.4.3			
Parent Template		none			
General Description		The Segmental Wall Analysis Entry is an Organizer with subsidiary entries that further describe the wall motion findings.			
Class/Mood	Code	Value Type	Value		
BATTERY / EVN	18118-0, LOINC, “LV Wall Motion Segmental Findings”				
Opt	entryRelationship	Description	Template ID	Spec Document	Con-straint
C [1..*]	COMP	Simple Observation (wall motion)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.4.4.1
C [1..*]	COMP	Simple Observation (wall morphology)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.4.4.2
O [0..1]	COMP	Simple Observation (viability)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.4.4.3
O [0..1]	COMP	observationMedia Entry	1.3.6.1.4.1.19376.1.4.1.4.7	CARD TF-2 6.2.4.6	CARD TF-2 6.2.4.4

The organizer Battery Act/statusCode value shall be 'completed'.

1455 For each of the values in the targetSite Value Set for which a wall motion assessment is made, there SHALL be a Simple Observation (wall motion) or a Simple Observation (wall morphology), or both.

There MAY be one ObservationMedia entry, whose value of type ED shall contain an in-line encoding of a graphic depiction of the wall analysis, e.g., a bull’s-eye diagram.

6.2.4.4.1 Simple Observation (wall motion) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
C [1..*]	R: LVG	60797005, SNOMED CT, “Cardiac Wall Motion” targetSiteCode from 1.2.840.10008.6.1.219 DICOM CID 3718 Myocardial Wall Segments in Projection	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.20 Wall motion
C [1..*]	R: SPECT, TTE, TEE, CMR O:CCTA	60797005, SNOMED CT, “Cardiac Wall Motion” targetSiteCode from 1.2.840.10008.6.1.218 DICOM CID 3717 Myocardial Wall Segments	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.20 Wall motion

1460

6.2.4.4.2 Simple Observation (wall morphology) Constraints

OPT	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
C [1..*]	R: Cath with LVG	72724002, SNOMED CT, “Morphology findings” targetSiteCode from 1.2.840.10008.6.1.219 DICOM CID 3718 Myocardial Wall Segments in Projection	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.19 Myocardium Assessments
C [1..*]	R: SPECT, echo, CMR O:CCTA	72724002, SNOMED CT, “Morphology findings” targetSiteCode from 1.2.840.10008.6.1.218 DICOM CID 3717 Myocardial Wall Segments	CD	n/a	1.3.6.1.4.1.19376.1.4.1.5.19 Myocardium Assessments

The observation/value MAY be a null flavor.

A morphological assessment observation MAY have a subsidiary Severity observation (templateID 1.3.6.1.4.1.19376.1.5.3.1.4.1 [PCC TF-2]).

1465

6.2.4.4.3 Simple Observation (viability) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
O [0..1]		404684003, SNOMED CT, “Finding”	ED (text/plain)		degree of viability in the infarct zone <ul style="list-style-type: none"> • Small • Moderate • Large • None

6.2.4.5 Coronary Angiography Arterial Analysis

Template ID	1.3.6.1.4.1.19376.1.4.1.4.5		
Parent Template	none		
General Description	The Coronary Angiography Arterial Analysis Entry is an Organizer with subsidiary entries that describe the findings for each arterial segment.		
Class/Mood	Code	Value Type	Value
BATTERY / EVN	33367005, SNOMED CT, “Coronary Angiography”		

Opt	entryRelationship	Description	Template ID	Spec Document	Constraint
R [1..*]	COMP	Simple Observation (stenosis)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.4.5.1
R [1..2]	COMP	Simple Observation (coronary findings)	1.3.6.1.4.1.19376.1.5.3.1.4.13	PCC TF-2	CARD TF-2 6.2.4.5.2
O [0..1]	COMP	observationMedia Entry	1.3.6.1.4.1.19376.1.4.1.4.7	CARD TF-2 6.2.4.6	CARD TF-2 6.2.4.5

1470 The organizer `Battery Act/statusCode` value shall be 'completed'.

There SHALL be one `ActRelationship` with `typeCode` COMP with a target of an observation for each arterial segment visualized.

1475 There SHOULD be one `ActRelationship` from the Organizer with `typeCode` COMP with a target of an `ObservationMedia`, whose value is a graphic depiction of the coronary analysis (e.g., a coronary tree diagram).

6.2.4.5.1 Simple Observation (stenosis) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
C [1..*]	R: ICA, CCTA	408715008, SNOMED CT, "Lumen Diameter Stenosis" targetSiteCode from 1.2.840.10008.6.1.48 DICOM CID 3015 Coronary Arteries	PQ or IVL<PQ>	%	

The Simple Observations Entry template provides additional guidance regarding observation elements `templateId`, `id`, `text`, `statusCode`, `effectiveTime`, and `text`.

1480 The stenosis observation/value MAY be a categorical assessment using the HL7v3 Interval of Physical Quantities data type (IVL<PQ>), with preferred value ranges as shown in Table 6.2.4.5-1.

Note: The observation/value may be a null flavor.

1485 **Table 6.2.4.5-1 Preferred stenosis categorical ranges**

Preferred stenosis categorical ranges
<10 %
10-50 %
50-70 %
70-90 %
>90 %

6.2.4.5.2 Simple Observation (coronary findings) Constraints

Opt	Exam Type Condition	observation/code	Data Type	Unit of Measure	Value Set
R [1..1]		404684003, SNOMED CT, "Finding"	CD		1.2.840.10008.6.1.211 DICOM CID 3710 Coronary Dominance
O [0..1]		414024009, SNOMED CT, "Disorder of coronary artery"	ED text/plain		indicate whether coronary anomalies, such as abnormal origin or location, are present

6.2.4.6 observationMedia Entry

Template ID	1.3.6.1.4.1.19376.1.4.1.4.7		
Parent Template	none		
General Description	The observationMedia Entry provides an in-line graphic depiction of the section findings. It is referenced by a <renderMultiMedia> element in the section text.		
Class/Mood	Code	Value Type	Value
OBS / EVN		ED	<tel> element SHALL NOT be present

1490

The value of type ED SHALL contain an in-line encoding of a graphic (i.e., the <tel> element shall not be present).

The ObservationMedia entry SHALL include an XML IDREFS ID attribute used as a target of a <renderMultiMedia> element in the section/text narrative block of the parent section.

1495

6.2.4.7 Planned Procedure Activity Entry

Template ID	1.3.6.1.4.1.19376.1.4.1.4.8				
Parent Template	none				
General Description	The Planned Procedure Activity Entry is a description of a planned activity, with subsidiary optional entries that further describe the reasons for the activity.				
Class/Mood	Code	Value Type	Value		
PROC / INT	unconstrained				
Opt	entryRelationship	Description	Template ID	Spec Document	Constraint
O [0..*]	RSON	Problem Entry	1.3.6.1.4.1.19376.1.5.3.1.4.5	PCC TF-2	

The Planned Procedure Activity entry SHALL describe the planned activity in either the act / code element, or in the act / text element.

1500 **6.3 Value Sets****1.3.6.1.4.1.19376.1.4.1.5.1 Cardiac imaging services**

Concept	Coding Scheme	SNOMED CT
Echocardiography		40701008
CCT (cardiac computed tomography)		241547009
CMR (cardiac magnetic resonance)		241620005
Cardiovascular NM		108294005
Cardiac PET (positron emission tomography)		241439007
coronary angiography and left ventriculography		418903008

1.3.6.1.4.1.19376.1.4.1.5.2 Cardiac imaging procedures

Concept	Coding Scheme	SNOMED CT	Abbreviation
radionuclide myocardial perfusion study		252432008	SPECT
PET heart study		241439007	PET
radionuclide angiocardiology		426940008	RNA
exercise stress echocardiography		433233004	Stress TTE
transthoracic echocardiography		433236007	TTE
transesophageal echocardiography		105376000	TEE
cardiac CT for calcium scoring		426005005	CACS
CT angiography of coronary arteries		419545005	CCTA
CACS and CCTA		426005005+419545005	
cardiac MRI		241620005	CMR
MRI stress study of cardiac function		431609005	Stress CMR
fluoroscopic angiography of coronary arteries		419416005	ICA
fluoroscopic angiography of left ventricle and coronary arteries		418903008	ICA+LVG

1505

1.3.6.1.4.1.19376.1.4.1.5.3 Referring Physician Specialty

Concept	Coding Scheme	nuccProviderCodes	SNOMED CT
Cardiologist		207RC0000X	17561000
Family practice		207Q00000X	62247001
Internal medicine		207R00000X	39677007
OB/GYN		207V00000X	309367003
Hospitalist		208M00000X	309395003

Concept	Coding Scheme	nuccProviderCodes	SNOMED CT
Surgeon		208600000X	304292004
Anesthesiologist		207L00000X	88189002
Radiologist		2085R0202X	66862007
Emergency department physician		207P00000X	309294001
Physician extender			

Note: As described in Section 6.1.2.4, the selection of the appropriate coding system for use may be based on local policy or national regulation.

1510 **1.3.6.1.4.1.19376.1.4.1.5.4 Cardiac problems/concerns**

Concept	Coding Scheme	SNOMED CT
Hypertension		38341003
Dyslipidemia		370992007
Diabetes		73211009
Acute renal failure		14669001
Chronic kidney disease		236425005
Peripheral arterial disease		399957001
Cerebrovascular disease		62914000
Erectile dysfunction		398175007
Cardiac arrhythmia		44808001
Asthma		195967001
Bronchospasm		4386001
Implanted pacemaker		371821000
Heart failure		84114007
Myocardial infarction		22298006
Angina		194828000

1.3.6.1.4.1.19376.1.4.1.5.5 Cardiac interventions

Concept	Coding Scheme	SNOMED CT
Percutaneous coronary intervention		415070008
Coronary artery bypass graft		232717009
Implantation of cardiac pacemaker		307280005

1.3.6.1.4.1.19376.1.4.1.5.7 Angina Type

Concept	Coding Scheme	SNOMED CT
Stable angina		233819005
Unstable angina		4557003
Atypical chest pain		371807002
Myocardial infarction		22298006

1515

1.3.6.1.4.1.19376.1.4.1.5.10 Contrast Agents Classes for Adverse Reactions

Concept	Coding Scheme	SNOMED CT
Iodinated contrast agent		426722004
Gadolinium compound		105879004
Echocardiography agent		409290009
radiopharmaceutical		349358000

1.3.6.1.4.1.19376.1.4.1.5.11 Contrast Agents

Concept	Coding Scheme	SNOMED CT	NDC
Radionuclide: F-18 FDG for viability		422975006	
Radionuclide: Rubidium-82 perfusion		79197006	
Radionuclide: Nitrogen-13 ammonia perfusion		21576001	
Radionuclide: Tc-99m tetrofosmin (Myoview)		404707004	
Radionuclide: Tc-99m sestamibi (Cardiolite)		404706008	
Radionuclide: Tl-201		353842007	
Echo Contrast: Optison (Perflutren)		409291008	00019-2707-03
Echo Contrast: Definity (Perflutren Lipid Microsphere)			11994-*011-04
Echo Contrast: Agitated saline		373757009	
Echo Contrast: Iodinated contrast		426722004	
High Osmolar Ionic Contrast: Diatrizoate meglumine and diatrizoate sodium (Renografin, etc.)		416688007	
High Osmolar Ionic Contrast: Iothalamate dimeglumine (Conray)		109221002	
Low osmolar non-ionic contrast: Iopamidol (Isovue)		109219007	
Low osmolar non-ionic contrast: Iohexol (Omnipaque)		109218004	
Low osmolar non-ionic contrast: Ioversol (Optiray)		109222009	
Low osmolar non-ionic contrast: Ioxaglate (Hexabrix)		353924001	
Low osmolar non-ionic contrast: Iomeprol (Iomeron)		356671000	
Low osmolar non-ionic contrast: Iopromide (Ultravist)		353903006	
Iso-osmolar nonionic contrast: Iodixanol (VisiPaque)		353962003	

Concept	Coding Scheme	SNOMED CT	NDC
Paramagnetic agent: Gadopentetate dimeglumine (Magnevist)		404846007	
Paramagnetic agent: Gadodiamide (Omniscan)		354088005	
Paramagnetic agent: Gadoversetamide (Optimark)		409477004	
Paramagnetic agent: Gadobenate dimeglumine (MultiHance)		414307008	

1520

1.3.6.1.4.1.19376.1.4.1.5.13 Cardiac Lab Results

Concept	Coding Scheme	LOINC
Cholesterol.in HDL		2085-9
Cholesterol.in LDL		2089-1
Cholesterol		2093-3
Triglyceride		2571-8
High sensitivity C reactive protein		30522-7
Creatine kinase.MB		13969-1
Natriuretic peptide.B		30934-4
Natriuretic peptide.B prohormone		33762-6
Troponin T.cardiac		6598-7
Creatinine		2160-0
Hemoglobin A1c		41995-2
Urea nitrogen		3094-0
Fasting glucose		1557-8

1.3.6.1.4.1.19376.1.4.1.5.14 Cardiac Drug Classes

Concept	Coding Scheme	SNOMED CT	NDF-RT
ACE inhibitor		69306018	N0000029130
Angiotensin receptor blocker		96308008	N0000175561
Aspirin, other antiplatelet agents		7947003	N0000145918
Calcium channel blockers		48698004	N0000029119
Beta-blockers		33252009	N0000029118
Erectile dysfunction medication		407315009	
Nitrates		31970009	N0000007647
Warfarin		48603004	N0000148057
Antiarrhythmics		67507000	N0000029121
Digitalis		65774009	N0000147198
Metformin		109081006	N0000021984

Concept	Coding Scheme	SNOMED CT	NDF-RT
Lipid-lowering medication		57952007	N0000029122
Other antihypertensives			N0000029427
Aminophylline		55867006	N0000146397
Theophylline		66493003	N0000146467
Dipyridamole		66859009	N0000146237
Inhaler			N0000177906
Diabetic medications		384953001	

Note: As described in Section 6.1.2.4, the selection of the appropriate coding system for use may be based on local policy or national regulation.

1525

1.3.6.1.4.1.19376.1.4.1.5.15 Drug Classes Used in Cardiac Procedure

Concept	Coding Scheme	SNOMED CT	NDF-RT
Calcium channel blockers		48698004	N0000029119
Beta-blockers		33252009	N0000029118
Nitrates		31970009	N0000007647
Aminophylline		55867006	N0000146397

Note: As described in Section 6.1.2.4, the selection of the appropriate coding system for use may be based on local policy or national regulation.

1530

1.3.6.1.4.1.19376.1.4.1.5.17 Rhythm Findings

Concept	Coding Scheme	SNOMED	MDC Code	MDC Ref
Normal sinus rhythm		64730000	10:9232	MDC_ECG_RHY_SINUS_NORMAL_RHY
Atrial fibrillation		49436004	10:9472	MDC_ECG_RHY_ATR_FIB
Premature atrial contractions		425880007	10:8272	MDC_ECG_BEAT_ATR_P_C
Premature ventricular contractions		427172004	10:8336	MDC_ECG_BEAT_V_P_C
Paced rhythm		426453001	10:8704	MDC_ECG_BEAT_PACED
Atrial flutter		164890007	10:9456	MDC_ECG_RHY_ATR_FLUT
Sinus tachycardia		427084000	10:9264	MDC_ECG_RHY_SINUS_TACHY
Sinus bradycardia		426177001	10:9248	MDC_ECG_RHY_SINUS_BRADY

Note: As described in Section 6.1.2.4, the selection of the appropriate coding system for use may be based on local policy or national regulation.

1.3.6.1.4.1.19376.1.4.1.5.18 ECG Findings

Concept	Coding Scheme	SNOMED	MDC Code	MDC Ref
---------	---------------	--------	----------	---------

Left bundle branch block	64730000	10:9232	MDC_ECG_BEAT_LBB_BLK_COMP
Right bundle branch block	49436004	10:9472	MDC_ECG_BEAT_RBB_BLK_COMP
Ventricular paced rhythm	61552005	10:5120	MDC_ECG_WAVP_PACE_V
Pre-excitation	195060002	10:10096	MDC_ECG_RHY_PREX

1535

Note: As described in Section 6.1.2.4, the selection of the appropriate coding system for use may be based on local policy or national regulation.

1.3.6.1.4.1.19376.1.4.1.5.19 Myocardium Assessments

Concept	Coding Scheme	SNOMED CT
normal wall motion		373122000
Scar		12402003
Myocardial ischemia		414795007
Myocardial infarction		22298006
Mixed myocardial ischemia and infarction		428196007
Mixed scar and ischemia		12402003+414795007
Ventricular aneurysm		90539001

1540

1.3.6.1.4.1.19376.1.4.1.5.20 Wall motion

Concept	Coding Scheme	SNOMED CT
Normal wall motion		373122000
Hyperkinetic region		373123005
Hypokinesis		37706002
Akinesis		195675009
Dyskinesis		25437005

1.3.6.1.4.1.19376.1.4.1.5.22 Cardiac Chamber Size Assessments

Concept	Coding Scheme	SNOMED CT
normal size cardiac chamber		373124004
abnormally small cardiac chamber		373125003
mildly enlarged cardiac chamber		373126002
moderately enlarged cardiac chamber		373127006
markedly enlarged cardiac chamber		373128001

1545

1.3.6.1.4.1.19376.1.4.1.5.23 Pulmonary Veins Assessments

Concept	Coding Scheme	SNOMED CT
pulmonary venous connections normal		446158009
variant number of pulmonary veins (usually 3 or 5), but with normal pulmonary venous drainage into left atrium		
anomalous pulmonary venous drainage		59631007

1.3.6.1.4.1.19376.1.4.1.5.24 Imaging Findings

Concept	Coding Scheme	SNOMED CT
Imaging result normal		408573005
Imaging result abnormal		408574004
Imaging result equivocal		408379005

1.3.6.1.4.1.19376.1.4.1.5.25 Angina Characteristics

Concept	Coding Scheme	SNOMED CT
Substernal chest pain		4568003
Provoked by exertion		427341007
Relieved by rest		427935006
Relieved by medication - nitroglycerin		428346000:363701004=71759000

1550

1.3.6.1.4.1.19376.1.4.1.5.26 Cardiac risk scores

Concept	Coding Scheme	LOINC
Hard coronary heart disease 10Y risk, Framingham		65850-0
Coronary heart disease 10Y risk, Framingham		65851-8
Coronary heart disease 2Y risk, Framingham		65852-6
General cardiovascular disease 10Y risk, Framingham		65853-4
Intermittent claudication 4Y risk, Framingham		65854-2
Recurring coronary heart disease 2Y risk, Framingham		65862-5
Stroke 10Y risk, Framingham		65855-9
Stroke after atrial fibrillation 5Y risk, Framingham		65864-1
Stroke or death after atrial fibrillation 5Y risk, Framingham		65865-8
Atrial fibrillation 10Y risk, Framingham		65856-7
Heart failure 4Y risk, Framingham		65863-3
Coronary heart disease 10Y risk, Reynolds		65857-5

Concept	Coding Scheme	LOINC
Coronary event 10Y risk, PROCAM QuickCheck		65858-3
Coronary event 10Y risk, PROCAM HealthCheck		65859-1
Cardiovascular disease 10Y risk, SCORE PC		65860-9
Cardiovascular disease 10Y risk, SCORE Quick		65861-7

1.3.6.1.4.1.19376.1.4.1.5.27 Cardiac Imaging Procedure Indications

Concept	Coding Scheme	SNOMED-CT
Chest Pain		29857009
Pre-operative		262068006
Coronary Artery Disease		53741008
Heart failure		84114007
Heart disease risk factors		171224000
Dyspnea		267036007
Post PTCA		373108000
History of CABG		399261000
Abnormal exercise tolerance test		165084003
Abnormal ECG		102594003
Arrhythmia		44808001
Angina pectoris		194828000
Hypertension		38341003
Palpitations		80313002
Supraventricular tachycardia		6456007
Syncope		271594007
History of Myocardial Infarction		399211009
Left bundle branch block		63467002
Valvular heart disease		368009
Occupational requirement		429060002
cardiogenic shock		89138009
ischemic heart disease		414545008
cardiac function test abnormal		165076002
heart transplant		32413006
heart disease - congenital		13213009
cardiomyopathy		85898001
heart disease		56265001
structural disorder of heart		128599005
Pericardial disease		55855009

1555

1.3.6.1.4.1.19376.1.4.1.5.28 Cardiac Synchronization Techniques

Concept	Coding Scheme	DCM
Real time acquisition		109080
Prospective gating		109081
Retrospective gating		109082
Paced		109083

1.3.6.1.4.1.19376.1.4.1.5.29 Cardiac Shunt Types

Concept	Coding Scheme	SNOMED CT
patent foramen ovale		204317008
atrial septal defect		70142008
ventricular septal defect		30288003
patent ductus arteriosus		83330001

1560

1.3.6.1.4.1.19376.1.4.1.5.30 ECG Summary Findings

Concept	Coding Scheme	SNOMED CT	DCM
ECG Ischemia		164861001	
ECG Abnormal		102594003	
ECG Equivocal		370359005	
ECG Normal		164854000	
ECG Nondiagnostic			122753

1.2.840.10008.6.1.48 DICOM CID 3015 - Coronary Arteries

Concept	Coding Scheme	SNOMED CT
Left Main Coronary Artery		3227004
Left Main Coronary Artery Ostium		76862008
Left Anterior Descending Coronary Artery		59438005
Proximal Left Anterior Descending Coronary Artery		68787002
Mid Left Anterior Descending Coronary Artery		91748002
Distal Left Anterior Descending Coronary Artery		36672000
Left Posterior Descending Artery		56322004

Concept	Coding Scheme	SNOMED CT
Left Posterior Descending Circumflex Coronary Artery		91760001
Left Posterolateral Circumflex Coronary Artery		57823005
Right Coronary Artery		13647002
Right Coronary Artery Ostium		56789007
Proximal Right Coronary Artery		91083009
Mid Right Coronary Artery		13647002+255562008
Distal Right Coronary Artery		41879009
Circumflex Coronary Artery		57396003
Proximal Circumflex Coronary Artery		52433000
Mid Circumflex Coronary Artery		91753007
Distal Circumflex Coronary Artery		6511003
Posterior Descending Right Coronary Artery		53655008
Intermediate Artery (Ramus)		244252004
Right posterior AV Coronary Artery		12800002
1st Diagonal Coronary Artery		91750005
1st Left Posterolateral Coronary Artery		91757008
1st Marginal Coronary Artery		91754001
1st Right posterolateral Coronary Artery		91761002
1st Septal Coronary Artery		244251006
2nd Diagonal Coronary Artery		91751009
2nd Left Posterolateral Coronary Artery		91758003
2nd Marginal Coronary Artery		91755000
2nd Right Posterolateral Coronary Artery		91762009
3rd Diagonal Coronary Artery		91752002
3rd Left Posterolateral Coronary Artery		91759006
3rd Marginal Coronary Artery		91756004
3rd Right posterolateral Coronary Artery		91763004
Marginal Right Coronary Artery		22765000
AV groove continuation of Circumflex Artery		75902001

Adapted from DICOM PS3.16-2009

1565

1.2.840.10008.6.1.57 DICOM CID 3261 - Stress Protocols

Concept	Coding Scheme	SNOMED-CT
Balke protocol		129097005

Concept	Coding Scheme	SNOMED-CT
Bruce protocol		129095002
Ellestad protocol		129098000
Modified Bruce protocol		129096001
Modified Naughton protocol		129102008
Naughton protocol		129101001
Pepper protocol		129100000
Ramp protocol		129099008
Exercise stress ECG test		46136006
Stress test using Bicycle Ergometer		26046004
Pharmacologic Stress protocol		424064009
Dipyridamole Stress protocol		422685009
Adenosine Stress protocol		424444005
Dobutamine Stress protocol		424225000
Pharmacologic and exercise stress test		428813002
Stress test using cardiac pacing		428685003

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.211 DICOM CID 3710 - Coronary Dominance

Concept	Coding Scheme	SNOMED CT
Left Coronary Dominance		253729004
Right Coronary Dominance		253728007
Balanced Coronary Dominance		253730009

Adapted from DICOM PS3.16-2009

1570

1.2.840.10008.6.1.218 DICOM CID 3717 - Myocardial Wall Segments

Concept	Coding Scheme	SNOMED CT
left ventricle basal anterior segment		264850008
left ventricle basal anteroseptal segment		396482007
left ventricle basal inferoseptal segment		396646008
left ventricle basal inferior segment		264846001
left ventricle basal inferolateral segment		396652009
left ventricle basal anterolateral segment		396654005

Concept	Coding Scheme	SNOMED CT
left ventricle mid anterior segment		264848000
left ventricle mid anteroseptal segment		396647004
left ventricle mid inferoseptal segment		396649001
left ventricle mid inferior segment		264847005
left ventricle mid inferolateral segment		396655006
left ventricle mid anterolateral segment		396656007
left ventricle apical anterior segment		264844003
left ventricle apical septal segment		264845002
left ventricle apical inferior segment		264849008
left ventricle apical lateral segment		264853005
apex of left ventricle		128564006

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.219 DICOM CID 3718 - Myocardial Wall Segments in Projection

1575

Concept	Coding Scheme	SNOMED CT
left ventricle basal anterior segment		264850008
myocardium of anterolateral region		73050001
myocardium of apex of heart		47962008
myocardium of diaphragmatic region		72542009
left ventricle basal inferior segment		264846001
left ventricle basal lateral segment		277631004
myocardium of posterolateral region		33272004
myocardium of inferolateral region		16239001
left ventricle apical septal segment		264845002
left ventricular basal septal segment		277630003
left ventricular posterobasal segment		408720008

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.223 DICOM CID 3722 - Diabetic Therapy

Concept	Coding Scheme	SNOMED CT
Dietary Treatment		284071006
Oral Treatment		170746002
Insulin Regime		225302006

Adapted from DICOM PS3.16-2009

1580

1.2.840.10008.6.1.225 DICOM CID 3724 - Smoking History

Concept	Coding Scheme	SNOMED CT
No History of Smoking		266919005
Current Smoker		77176002
Former Smoker		8517006

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.227 DICOM CID 3727 - Indications for Catheterization

Concept	Coding Scheme	SNOMED CT
cardiogenic shock		89138009
valvular heart disease		368009
Arrhythmia		44808001
ischemic heart disease		414545008
cardiac function test abnormal		165076002
heart transplant		32413006
heart disease - congenital		13213009
cardiomyopathy		85898001
heart disease		56265001

1585

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.253 DICOM CID 3758 - Cardiovascular Family History

Concept	Coding Scheme	SNOMED CT
Family history of coronary artery disease		430091005
Family history: Diabetes mellitus		160303001
Family history of myocardial infarction		266897007
No Family history of Diabetes		160274005
No Family history of Cardiovascular disease		160270001
Family History Unknown		407559004

Adapted from DICOM PS3.16-2009

1590

1.2.840.10008.6.1.616 DICOM CID 12224 - Ultrasound Image Modes

Concept	Coding Scheme	SNOMED CT
2D mode		399064001
Doppler Color Flow		261197005
M mode		399155008
Doppler Pulsed		261199008
Doppler Continuous Wave		261198000
Power Doppler		425704008
3D mode		426865009

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.744 DICOM CID 3112 - Attenuation Correction

Concept	Coding Scheme	DCM
Algorithmic Attenuation Correction		122726
NM Transmission Attenuation Correction		122727
CT-based Attenuation Correction		122728
No Attenuation Correction		122729

Adapted from DICOM PS3.16-2009

1595

1.2.840.10008.6.1.755 DICOM CID 3200 - Stress Test Procedure

Concept	Coding Scheme	SNOMED-CT
Exercise stress test		165079009
Pharmacologic stress test		424064009
Pharmacologic and exercise stress test		428813002
Paced stress test		428685003

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.759 DICOM CID 3204 - Stress Agents

Concept	Coding Scheme	SNOMED-CT
Dipyridamole		66859009
Dobutamine		26523005

Concept	Coding Scheme	SNOMED-CT
Adenosine		108502004
Atropine		73949004
Adenosine A2 receptor agonist		432062000

1600

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.770 DICOM CID 3217 - Comparison Finding

Concept	Coding Scheme	DCM
Agreement with prior findings		122775
Disagreement with prior findings		122776

Adapted from DICOM PS3.16-2009

1605

1.2.840.10008.6.1.781 DICOM CID 3235 - Perfusion Comparison Findings

Concept	Coding Scheme	SNOMED-CT
No change		260388006
New ischemia		428927006
Less ischemia		429232006
Resolution of ischemia		428824000
More ischemia		429477006
New infarction		429391004

Adapted from DICOM PS3.16-2009

1.2.840.10008.6.1.782 DICOM CID 3236 - Tolerance Comparison Findings

Concept	Coding Scheme	SNOMED-CT
No change		260388006
Decreased tolerance		102460003
Increased tolerance		102459008

Adapted from DICOM PS3.16-2009

1610

1.2.840.10008.6.1.783 DICOM CID 3237 - Wall Motion Comparison Findings

Concept	Coding Scheme	SNOMED-CT
No change		260388006
New wall motion abnormality		429058004
Improvement of wall motion		428825004

Adapted from DICOM PS3.16-2009

2.16.840.1.113883.3.88.12.3221.5.2 HITSP Health Insurance Type

1615 This value set uses the ASC X12 vocabulary for Insurance Type Code (ASC X12 Data Element 1336) and identifies the type of health plan covering the individual (e.g., an HMO, PPO, POS, etc.). The use of this value set is limited to the United States.

Concept	Coding Scheme	X12DE1336
Medicare Secondary Working Aged Beneficiary or Spouse with Employer Group Health Plan		12
Medicare Secondary End-Stage Renal Disease Beneficiary in the 12 month coordination period with an employer's group health plan		13
Medicare Secondary, No-fault Insurance including Auto is Primary		14
Medicare Secondary Worker's Compensation		15
Medicare Secondary Public Health Service (PHS) or Other Federal Agency		16
Medicare Secondary Black Lung		41
Medicare Secondary Veteran's Administration		42
Medicare Secondary Disabled Beneficiary Under Age 65 with Large Group Health Plan (LGHP)		43
Medicare Secondary, Other Liability Insurance is Primary		47
Auto Insurance Policy		AP
Commercial		C1
Consolidated Omnibus Budget Reconciliation Act (COBRA)		CO
Medicare Conditionally Primary		CP
Disability		D
Disability Benefits		DB
Exclusive Provider Organization		EP
Family or Friends		FF
Group Policy		GP
Health Maintenance Organization (HMO)		HM
Health Maintenance Organization (HMO) - Medicare Risk		HN

Concept	Coding Scheme	X12DE1336
Special Low Income Medicare Beneficiary		HS
Indemnity		IN
Individual Policy		IP
Long Term Care		LC
Long Term Policy		LD
Life Insurance		LI
Litigation		LT
Medicare Part A		MA
Medicare Part B		MB
Medicaid		MC
Medigap Part A		MH
Medigap Part B		MI
Medicare Primary		MP
Other		OT
Property Insurance – Personal		PE
Personal		PL
Personal Payment (Cash - No Insurance)		PP
Preferred Provider Organization (PPO)		PR
Point of Service (POS)		PS
Qualified Medicare Beneficiary		QM
Property Insurance – Real		RP
Supplemental Policy		SP
Tax Equity Fiscal Responsibility Act (TEFRA)		TF
Workers Compensation		WC
Wrap Up Policy		WU

Adapted from HITSP C80 Clinical Document and Message Terminology Component V2.0.1

1620